

Data Path : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\FID_B\Data\FB122618\
 Data File : FB019203.D
 Signal(s) : FID2B.CH
 Acq On : 26 Dec 2018 21:42
 Operator : MJ/SJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :

Integration File: Calibration.e
 Quant Time: Dec 27 09:24:29 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB120718.M
 Quant Title :
 QLast Update : Fri Dec 07 17:57:38 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.103	171553	19.627 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.957	409951	26.816 ng/ml
2) t 2,2,4-Trimethylpentane	7.719	446990	29.014 ng/ml
3) t n-Heptane	8.059	144308	9.709 ng/ml
4) t Benzene	8.197	181322	10.067 ng/ml
6) t Toluene	10.952	477643	31.467 ng/ml
7) t Ethylbenzene	13.401	159496	10.266 ng/ml
8) t m-Xylene	13.534	342038	20.995 ng/ml
9) t O-Xylene	14.265	358180	21.184 ng/ml
10) t 1,2,4-Trimethylbenzene	16.549	412091	20.363 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : \\TERASTORAGE\Terastorage\pestpcbsrv\HPCHEM1\FID_B\Data\FB122618\
Data File : FB019203.D
Signal(s) : FID2B.CH
Acq On : 26 Dec 2018 21:42
Operator : MJ/SJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :

Integration File: Calibration.e
Quant Time: Dec 27 09:24:29 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB120718.M
Quant Title :
QLast Update : Fri Dec 07 17:57:38 2018
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.957 min
Delta R.T.: -0.004 min
Response: 409951
Conc: 26.82 ng/ml

Instrument :
FID_B
ClientSampleId :




#2 2,2,4-Trimethylpentane

R.T.: 7.719 min
Delta R.T.: -0.004 min
Response: 446990
Conc: 29.01 ng/ml




#3 n-Heptane

R.T.: 8.059 min
Delta R.T.: -0.003 min
Response: 144308
Conc: 9.71 ng/ml



#4 Benzene


R.T.: 8.197 min
Delta R.T.: -0.004 min
Response: 181322
Conc: 10.07 ng/ml



#5 AAA-TFT


R.T.: 9.103 min
Delta R.T.: -0.004 min
Response: 171553
Conc: 19.63 ng/ml

Instrument :
FID_B
ClientSampleId :




#6 Toluene

R.T.: 10.952 min
Delta R.T.: -0.004 min
Response: 477643
Conc: 31.47 ng/ml




#7 Ethylbenzene

R.T.: 13.401 min
Delta R.T.: -0.004 min
Response: 159496
Conc: 10.27 ng/ml



#8 m-Xylene


R.T.: 13.534 min
Delta R.T.: -0.004 min
Response: 342038
Conc: 20.99 ng/ml



#9 O-Xylene

R.T.: 14.265 min
Delta R.T.: -0.004 min
Response: 358180
Conc: 21.18 ng/ml

Instrument :
FID_B
ClientSampleId :



#10 1,2,4-Trimethylbenzene

R.T.: 16.549 min
Delta R.T.: -0.004 min
Response: 412091
Conc: 20.36 ng/ml