

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB122619\
 Data File : FB023311.D
 Signal(s) : FID2B.CH
 Acq On : 26 Dec 2019 14:32
 Operator : DD/AJ
 Sample : K6118-11MSD
 Misc : 5.00g/5.00ml DI WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :
 SU-01-122319MSD

Integration File: Calibration.e
 Quant Time: Dec 27 00:31:48 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB121919.M
 Quant Title :
 QLast Update : Thu Dec 19 12:43:06 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.795	409672	17.276 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.686	828958	26.784 ng/ml
2) t 2,2,4-Trimethylpentane	7.413	892317	26.711 ng/ml
3) t n-Heptane	7.754	289265	8.769 ng/ml
4) t Benzene	7.888	421933	10.566 ng/ml
6) t Toluene	10.626	1151848	29.915 ng/ml
7) t Ethylbenzene	13.066	369889	10.191 ng/ml
8) t m-Xylene	13.199	679369	19.572 ng/ml
9) t O-Xylene	13.927	671431	19.934 ng/ml
10) t 1,2,4-Trimethylbenzene	16.205	546196	19.467 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB122619\
Data File : FB023311.D
Signal(s) : FID2B.CH
Acq On : 26 Dec 2019 14:32
Operator : DD/AJ
Sample : K6118-11MSD
Misc : 5.00g/5.00ml DI WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
SU-01-122319MSD

Integration File: Calibration.e
Quant Time: Dec 27 00:31:48 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB121919.M
Quant Title :
QLast Update : Thu Dec 19 12:43:06 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane


R.T.: 4.686 min
Delta R.T.: 0.002 min
Response: 828958
Conc: 26.78 ng/ml

Instrument :
FID_B
ClientSampleId :
SU-01-122319MSD




#2 2,2,4-Trimethylpentane

R.T.: 7.413 min
Delta R.T.: 0.002 min
Response: 892317
Conc: 26.71 ng/ml




#3 n-Heptane

R.T.: 7.754 min
Delta R.T.: 0.002 min
Response: 289265
Conc: 8.77 ng/ml



#4 Benzene


R.T.: 7.888 min
Delta R.T.: 0.002 min
Response: 421933
Conc: 10.57 ng/ml



#5 AAA-TFT


R.T.: 8.795 min
Delta R.T.: 0.002 min
Response: 409672
Conc: 17.28 ng/ml

Instrument :
FID_B
ClientSampleId :
SU-01-122319MSD




#6 Toluene

R.T.: 10.626 min
Delta R.T.: 0.002 min
Response: 1151848
Conc: 29.92 ng/ml




#7 Ethylbenzene

R.T.: 13.066 min
Delta R.T.: 0.002 min
Response: 369889
Conc: 10.19 ng/ml



#8 m-Xylene


R.T.: 13.199 min
Delta R.T.: 0.002 min
Response: 679369
Conc: 19.57 ng/ml



#9 O-Xylene

R.T.: 13.927 min
Delta R.T.: 0.002 min
Response: 671431
Conc: 19.93 ng/ml

Instrument :
FID_B
ClientSampleId :
SU-01-122319MSD



#10 1,2,4-Trimethylbenzene

R.T.: 16.205 min
Delta R.T.: 0.002 min
Response: 546196
Conc: 19.47 ng/ml