

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB122723\
 Data File : FB030515.D
 Signal(s) : FID2B.CH
 Acq On : 27 Dec 2023 10:59
 Operator : YP/AJ
 Sample : 20 GRO STD
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Dec 27 10:03:53 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB122723.M
 Quant Title :
 QLast Update : Wed Dec 27 10:03:23 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.785	230328	20.000 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.707	467252	30.000 ng/ml
2) t 2,2,4-Trimethylpentane	7.410	518319	30.000 ng/ml
3) t n-Heptane	7.745	161621	10.000 ng/ml
4) t Benzene	7.881	201903	10.000 ng/ml
6) t Toluene	10.612	589652	30.000 ng/ml
7) t Ethylbenzene	13.048	190885	10.000 ng/ml
8) t m-Xylene	13.181	383666	20.000 ng/ml
9) t O-Xylene	13.908	382438	20.000 ng/ml
10) t 1,2,4-Trimethylbenzene	16.183	402780	20.000 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB122723\
Data File : FB030515.D
Signal(s) : FID2B.CH
Acq On : 27 Dec 2023 10:59
Operator : YP/AJ
Sample : 20 GRO STD
Misc :
ALS Vial : 3 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Dec 27 10:03:53 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB122723.M
Quant Title :
QLast Update : Wed Dec 27 10:03:23 2023
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.707 min
Delta R.T.: 0.000 min
Response: 467252
Conc: 30.00 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.410 min
Delta R.T.: 0.000 min
Response: 518319
Conc: 30.00 ng/ml




#3 n-Heptane

R.T.: 7.745 min
Delta R.T.: 0.000 min
Response: 161621
Conc: 10.00 ng/ml




#4 Benzene

R.T.: 7.881 min
Delta R.T.: 0.000 min
Response: 201903
Conc: 10.00 ng/ml




#5 AAA-TFT

R.T.: 8.785 min
Delta R.T.: 0.000 min
Response: 230328
Conc: 20.00 ng/ml




#6 Toluene

R.T.: 10.612 min
Delta R.T.: 0.000 min
Response: 589652
Conc: 30.00 ng/ml




#7 Ethylbenzene

R.T.: 13.048 min
Delta R.T.: 0.000 min
Response: 190885
Conc: 10.00 ng/ml



#8 m-Xylene

R.T.: 13.181 min
Delta R.T.: 0.000 min
Response: 383666
Conc: 20.00 ng/ml




#9 O-Xylene

R.T.: 13.908 min

Delta R.T.: 0.000 min

Response: 382438

Conc: 20.00 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.183 min

Delta R.T.: 0.000 min

Response: 402780

Conc: 20.00 ng/ml