

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB123019\
 Data File : FB023344.D
 Signal(s) : FID2B.CH
 Acq On : 30 Dec 2019 15:52
 Operator : DD/AJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Dec 31 00:38:52 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB121919.M
 Quant Title :
 QLast Update : Thu Dec 19 12:43:06 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.797	473716	19.976 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.688	936978	30.274 ng/ml
2) t 2,2,4-Trimethylpentane	7.416	1047594	31.360 ng/ml
3) t n-Heptane	7.756	341675	10.358 ng/ml
4) t Benzene	7.890	407337	10.201 ng/ml
6) t Toluene	10.629	1174099	30.493 ng/ml
7) t Ethylbenzene	13.070	360473	9.931 ng/ml
8) t m-Xylene	13.203	691724	19.927 ng/ml
9) t O-Xylene	13.931	669598	19.880 ng/ml
10) t 1,2,4-Trimethylbenzene	16.209	576820	20.558 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB123019\
Data File : FB023344.D
Signal(s) : FID2B.CH
Acq On : 30 Dec 2019 15:52
Operator : DD/AJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 8 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Dec 31 00:38:52 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB121919.M
Quant Title :
QLast Update : Thu Dec 19 12:43:06 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.688 min
Delta R.T.: 0.004 min
Response: 936978
Conc: 30.27 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.416 min
Delta R.T.: 0.004 min
Response: 1047594
Conc: 31.36 ng/ml




#3 n-Heptane

R.T.: 7.756 min
Delta R.T.: 0.004 min
Response: 341675
Conc: 10.36 ng/ml




#4 Benzene

R.T.: 7.890 min
Delta R.T.: 0.004 min
Response: 407337
Conc: 10.20 ng/ml




#5 AAA-TFT

R.T.: 8.797 min
Delta R.T.: 0.004 min
Response: 473716
Conc: 19.98 ng/ml




#6 Toluene

R.T.: 10.629 min
Delta R.T.: 0.005 min
Response: 1174099
Conc: 30.49 ng/ml




#7 Ethylbenzene

R.T.: 13.070 min
Delta R.T.: 0.006 min
Response: 360473
Conc: 9.93 ng/ml




#8 m-Xylene

R.T.: 13.203 min
Delta R.T.: 0.005 min
Response: 691724
Conc: 19.93 ng/ml



#9 O-Xylene

R.T.: 13.931 min
Delta R.T.: 0.006 min
Response: 669598
Conc: 19.88 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.209 min
Delta R.T.: 0.005 min
Response: 576820
Conc: 20.56 ng/ml