

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB123119\
 Data File : FB023347.D
 Signal(s) : FID2B.CH
 Acq On : 31 Dec 2019 9:37
 Operator : DD/AJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Jan 01 00:15:56 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB121919.M
 Quant Title :
 QLast Update : Thu Dec 19 12:43:06 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.793	513792	21.666 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.686	960221	31.025 ng/ml
2) t 2,2,4-Trimethylpentane	7.411	1063197	31.827 ng/ml
3) t n-Heptane	7.751	350559	10.628 ng/ml
4) t Benzene	7.886	422764	10.587 ng/ml
6) t Toluene	10.625	1240900	32.228 ng/ml
7) t Ethylbenzene	13.065	385472	10.620 ng/ml
8) t m-Xylene	13.198	738140	21.265 ng/ml
9) t O-Xylene	13.926	709750	21.072 ng/ml
10) t 1,2,4-Trimethylbenzene	16.204	600012	21.385 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB123119\
Data File : FB023347.D
Signal(s) : FID2B.CH
Acq On : 31 Dec 2019 9:37
Operator : DD/AJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 1 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Jan 01 00:15:56 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB121919.M
Quant Title :
QLast Update : Thu Dec 19 12:43:06 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.686 min
Delta R.T.: 0.002 min
Response: 960221
Conc: 31.03 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.411 min
Delta R.T.: 0.000 min
Response: 1063197
Conc: 31.83 ng/ml




#3 n-Heptane

R.T.: 7.751 min
Delta R.T.: 0.000 min
Response: 350559
Conc: 10.63 ng/ml




#4 Benzene

R.T.: 7.886 min
Delta R.T.: 0.000 min
Response: 422764
Conc: 10.59 ng/ml




#5 AAA-TFT

R.T.: 8.793 min
Delta R.T.: 0.000 min
Response: 513792
Conc: 21.67 ng/ml




#6 Toluene

R.T.: 10.625 min
Delta R.T.: 0.000 min
Response: 1240900
Conc: 32.23 ng/ml




#7 Ethylbenzene

R.T.: 13.065 min
Delta R.T.: 0.001 min
Response: 385472
Conc: 10.62 ng/ml




#8 m-Xylene

R.T.: 13.198 min
Delta R.T.: 0.000 min
Response: 738140
Conc: 21.26 ng/ml



#9 O-Xylene

R.T.: 13.926 min
Delta R.T.: 0.001 min
Response: 709750
Conc: 21.07 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.204 min
Delta R.T.: 0.001 min
Response: 600012
Conc: 21.38 ng/ml