

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB062618\
 Data File : FB016400.D
 Signal(s) : FID2B.CH
 Acq On : 26 Jun 2018 19:50
 Operator : MJ/SJ
 Sample : J3240-05MS
 Misc : 5.01g/5.00mL DI Water
 ALS Vial : 5 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Jun 27 09:34:34 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB060118.M
 Quant Title :
 QLast Update : Fri Jun 01 13:25:03 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.116	151074	21.497 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.968	304111	25.079 ng/ml
2) t 2,2,4-Trimethylpentane	7.732	316365	27.318 ng/ml
3) t n-Heptane	8.074	83535	7.633 ng/ml
4) t Benzene	8.211	159689	10.755 ng/ml
6) t Toluene	10.965	327511	26.385 ng/ml
7) t Ethylbenzene	13.414	128674	9.193 ng/ml
8) t m-Xylene	13.547	275318	18.088 ng/ml
9) t O-Xylene	14.278	297990	19.095 ng/ml
10) t 1,2,4-Trimethylbenzene	16.561	326360	18.949 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB062618\
Data File : FB016400.D
Signal(s) : FID2B.CH
Acq On : 26 Jun 2018 19:50
Operator : MJ/SJ
Sample : J3240-05MS
Misc : 5.01g/5.00mL DI Water
ALS Vial : 5 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Jun 27 09:34:34 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB060118.M
Quant Title :
QLast Update : Fri Jun 01 13:25:03 2018
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.968 min
Delta R.T.: 0.000 min
Response: 304111
Conc: 25.08 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.732 min
Delta R.T.: -0.002 min
Response: 316365
Conc: 27.32 ng/ml




#3 n-Heptane

R.T.: 8.074 min
Delta R.T.: 0.000 min
Response: 83535
Conc: 7.63 ng/ml




#4 Benzene

R.T.: 8.211 min
Delta R.T.: 0.000 min
Response: 159689
Conc: 10.75 ng/ml




#5 AAA-TFT

R.T.: 9.116 min
Delta R.T.: 0.000 min
Response: 151074
Conc: 21.50 ng/ml




#6 Toluene

R.T.: 10.965 min
Delta R.T.: 0.000 min
Response: 327511
Conc: 26.39 ng/ml




#7 Ethylbenzene

R.T.: 13.414 min
Delta R.T.: 0.000 min
Response: 128674
Conc: 9.19 ng/ml




#8 m-Xylene

R.T.: 13.547 min
Delta R.T.: 0.000 min
Response: 275318
Conc: 18.09 ng/ml



#9 O-Xylene

R.T.: 14.278 min
Delta R.T.: 0.000 min
Response: 297990
Conc: 19.10 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.561 min
Delta R.T.: 0.000 min
Response: 326360
Conc: 18.95 ng/ml