

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB082219\
Data File : FB022083.D
Signal(s) : FID2B.CH
Acq On : 22 Aug 2019 17:32
Operator : DD/AJ
Sample : K4462-04
Misc : 5.03/5.00ml DI WATER
ALS Vial : 14 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Aug 23 02:12:40 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB082019.M
Quant Title :
QLast Update : Wed Aug 21 07:59:08 2019
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.988	371736	17.502 ng/ml
Target Compounds			
6) t Toluene	10.841	3107	0.101 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB082219\
Data File : FB022083.D
Signal(s) : FID2B.CH
Acq On : 22 Aug 2019 17:32
Operator : DD/AJ
Sample : K4462-04
Misc : 5.03/5.00ml DI WATER
ALS Vial : 14 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Aug 23 02:12:40 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB082019.M
Quant Title :
QLast Update : Wed Aug 21 07:59:08 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um





#5 AAA-TFT

R.T.: 8.988 min
Delta R.T.: 0.006 min
Response: 371736
Conc: 17.50 ng/ml



#6 Toluene

R.T.: 10.841 min
Delta R.T.: 0.018 min
Response: 3107
Conc: 0.10 ng/ml