

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB082219\
 Data File : FB022094.D
 Signal(s) : FID2B.CH
 Acq On : 22 Aug 2019 23:13
 Operator : DD/AJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Aug 23 07:35:16 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB082019.M
 Quant Title :
 QLast Update : Wed Aug 21 07:59:08 2019
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.989	303947	14.310 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.859	838511	20.951 ng/ml
2) t 2,2,4-Trimethylpentane	7.609	856519	21.265 ng/ml
3) t n-Heptane	7.949	265667	6.955 ng/ml
4) t Benzene	8.084	366875	8.172 ng/ml
6) t Toluene	10.829	741666	24.125 ng/ml
7) t Ethylbenzene	13.275	170980	8.379 ng/ml
8) t m-Xylene	13.408	356214	16.914 ng/ml
9) t O-Xylene	14.137	395625	17.689 ng/ml
10) t 1,2,4-Trimethylbenzene	16.416	432766	16.875 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB082219\
Data File : FB022094.D
Signal(s) : FID2B.CH
Acq On : 22 Aug 2019 23:13
Operator : DD/AJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 25 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Aug 23 07:35:16 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB082019.M
Quant Title :
QLast Update : Wed Aug 21 07:59:08 2019
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.859 min
Delta R.T.: 0.007 min
Response: 838511
Conc: 20.95 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.609 min
Delta R.T.: 0.009 min
Response: 856519
Conc: 21.27 ng/ml




#3 n-Heptane

R.T.: 7.949 min
Delta R.T.: 0.008 min
Response: 265667
Conc: 6.95 ng/ml




#4 Benzene

R.T.: 8.084 min
Delta R.T.: 0.007 min
Response: 366875
Conc: 8.17 ng/ml




#5 AAA-TFT

R.T.: 8.989 min
Delta R.T.: 0.007 min
Response: 303947
Conc: 14.31 ng/ml




#6 Toluene

R.T.: 10.829 min
Delta R.T.: 0.007 min
Response: 741666
Conc: 24.13 ng/ml




#7 Ethylbenzene

R.T.: 13.275 min
Delta R.T.: 0.005 min
Response: 170980
Conc: 8.38 ng/ml




#8 m-Xylene

R.T.: 13.408 min
Delta R.T.: 0.006 min
Response: 356214
Conc: 16.91 ng/ml



#9 O-Xylene

R.T.: 14.137 min
Delta R.T.: 0.005 min
Response: 395625
Conc: 17.69 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.416 min
Delta R.T.: 0.004 min
Response: 432766
Conc: 16.87 ng/ml