

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB031320\
 Data File : FB023932.D
 Signal(s) : FID2B.CH
 Acq On : 13 Mar 2020 15:27
 Operator : DD/AJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Mar 13 22:38:52 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB030520.M
 Quant Title :
 QLast Update : Thu Mar 05 11:03:12 2020
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.803	413027	19.621 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.692	804134	26.020 ng/ml
2) t 2,2,4-Trimethylpentane	7.419	852338	25.276 ng/ml
3) t n-Heptane	7.761	276287	8.340 ng/ml
4) t Benzene	7.895	393888	9.407 ng/ml
6) t Toluene	10.634	1156237	28.358 ng/ml
7) t Ethylbenzene	13.074	362126	9.478 ng/ml
8) t m-Xylene	13.207	702321	19.076 ng/ml
9) t O-Xylene	13.935	676674	19.026 ng/ml
10) t 1,2,4-Trimethylbenzene	16.213	591335	19.519 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB031320\
Data File : FB023932.D
Signal(s) : FID2B.CH
Acq On : 13 Mar 2020 15:27
Operator : DD/AJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 13 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Mar 13 22:38:52 2020
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB030520.M
Quant Title :
QLast Update : Thu Mar 05 11:03:12 2020
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.692 min
Delta R.T.: 0.006 min
Response: 804134
Conc: 26.02 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.419 min
Delta R.T.: 0.006 min
Response: 852338
Conc: 25.28 ng/ml




#3 n-Heptane

R.T.: 7.761 min
Delta R.T.: 0.007 min
Response: 276287
Conc: 8.34 ng/ml




#4 Benzene

R.T.: 7.895 min
Delta R.T.: 0.007 min
Response: 393888
Conc: 9.41 ng/ml




#5 AAA-TFT

R.T.: 8.803 min
Delta R.T.: 0.008 min
Response: 413027
Conc: 19.62 ng/ml




#6 Toluene

R.T.: 10.634 min
Delta R.T.: 0.008 min
Response: 1156237
Conc: 28.36 ng/ml




#7 Ethylbenzene

R.T.: 13.074 min
Delta R.T.: 0.007 min
Response: 362126
Conc: 9.48 ng/ml




#8 m-Xylene

R.T.: 13.207 min
Delta R.T.: 0.007 min
Response: 702321
Conc: 19.08 ng/ml



#9 O-Xylene

R.T.: 13.935 min
Delta R.T.: 0.007 min
Response: 676674
Conc: 19.03 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.213 min
Delta R.T.: 0.007 min
Response: 591335
Conc: 19.52 ng/ml