

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB070218\
 Data File : FB016538.D
 Signal(s) : FID2B.CH
 Acq On : 3 Jul 2018 3:50
 Operator : MJ/SJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Jul 03 07:34:24 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB070218.M
 Quant Title :
 QLast Update : Mon Jul 02 14:53:33 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.116	147425	17.855 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.969	405917	28.654 ng/ml
2) t 2,2,4-Trimethylpentane	7.734	394578	28.656 ng/ml
3) t n-Heptane	8.072	123694	9.144 ng/ml
4) t Benzene	8.212	165002	8.823 ng/ml
6) t Toluene	10.966	376228	28.067 ng/ml
7) t Ethylbenzene	13.415	142044	9.728 ng/ml
8) t m-Xylene	13.547	293443	18.476 ng/ml
9) t O-Xylene	14.279	313037	18.334 ng/ml
10) t 1,2,4-Trimethylbenzene	16.560	312964	16.417 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB070218\
Data File : FB016538.D
Signal(s) : FID2B.CH
Acq On : 3 Jul 2018 3:50
Operator : MJ/SJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 33 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Jul 03 07:34:24 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB070218.M
Quant Title :
QLast Update : Mon Jul 02 14:53:33 2018
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.969 min
Delta R.T.: 0.003 min
Response: 405917
Conc: 28.65 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.734 min
Delta R.T.: 0.003 min
Response: 394578
Conc: 28.66 ng/ml




#3 n-Heptane

R.T.: 8.072 min
Delta R.T.: 0.002 min
Response: 123694
Conc: 9.14 ng/ml




#4 Benzene

R.T.: 8.212 min
Delta R.T.: 0.000 min
Response: 165002
Conc: 8.82 ng/ml




#5 AAA-TFT

R.T.: 9.116 min
Delta R.T.: 0.000 min
Response: 147425
Conc: 17.86 ng/ml




#6 Toluene

R.T.: 10.966 min
Delta R.T.: 0.000 min
Response: 376228
Conc: 28.07 ng/ml




#7 Ethylbenzene

R.T.: 13.415 min
Delta R.T.: 0.000 min
Response: 142044
Conc: 9.73 ng/ml




#8 m-Xylene

R.T.: 13.547 min
Delta R.T.: 0.000 min
Response: 293443
Conc: 18.48 ng/ml



#9 O-Xylene

R.T.: 14.279 min
Delta R.T.: 0.000 min
Response: 313037
Conc: 18.33 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.560 min
Delta R.T.: -0.001 min
Response: 312964
Conc: 16.42 ng/ml