

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB072718\
 Data File : FB016980.D
 Signal(s) : FID2B.CH
 Acq On : 27 Jul 2018 20:38
 Operator : MJ/SJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Jul 30 08:55:42 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB072718.M
 Quant Title :
 QLast Update : Fri Jul 27 13:23:50 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.114	167169	21.012 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.965	443616	31.183 ng/ml
2) t 2,2,4-Trimethylpentane	7.731	437137	31.867 ng/ml
3) t n-Heptane	8.069	139621	9.998 ng/ml
4) t Benzene	8.209	178905	10.214 ng/ml
6) t Toluene	10.963	425753	31.384 ng/ml
7) t Ethylbenzene	13.413	135831	9.898 ng/ml
8) t m-Xylene	13.546	290432	20.020 ng/ml
9) t O-Xylene	14.278	308454	20.049 ng/ml
10) t 1,2,4-Trimethylbenzene	16.559	365000	19.570 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB072718\
Data File : FB016980.D
Signal(s) : FID2B.CH
Acq On : 27 Jul 2018 20:38
Operator : MJ/SJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 17 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Jul 30 08:55:42 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB072718.M
Quant Title :
QLast Update : Fri Jul 27 13:23:50 2018
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.965 min
Delta R.T.: 0.001 min
Response: 443616
Conc: 31.18 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.731 min
Delta R.T.: 0.000 min
Response: 437137
Conc: 31.87 ng/ml




#3 n-Heptane

R.T.: 8.069 min
Delta R.T.: 0.000 min
Response: 139621
Conc: 10.00 ng/ml




#4 Benzene

R.T.: 8.209 min
Delta R.T.: 0.000 min
Response: 178905
Conc: 10.21 ng/ml




#5 AAA-TFT

R.T.: 9.114 min
Delta R.T.: 0.000 min
Response: 167169
Conc: 21.01 ng/ml




#6 Toluene

R.T.: 10.963 min
Delta R.T.: 0.000 min
Response: 425753
Conc: 31.38 ng/ml




#7 Ethylbenzene

R.T.: 13.413 min
Delta R.T.: 0.002 min
Response: 135831
Conc: 9.90 ng/ml




#8 m-Xylene

R.T.: 13.546 min
Delta R.T.: 0.002 min
Response: 290432
Conc: 20.02 ng/ml



#9 O-Xylene

R.T.: 14.278 min
Delta R.T.: 0.002 min
Response: 308454
Conc: 20.05 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.559 min
Delta R.T.: 0.001 min
Response: 365000
Conc: 19.57 ng/ml