

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB081418\
 Data File : FB017390.D
 Signal(s) : FID2B.CH
 Acq On : 15 Aug 2018 13:44
 Operator : MJ/SJ
 Sample : 20 PPB GRO STD
 Misc :
 ALS Vial : 43 Sample Multiplier: 1

Integration File: Calibration.e
 Quant Time: Aug 15 14:07:45 2018
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB080818.M
 Quant Title :
 QLast Update : Wed Aug 08 11:39:03 2018
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	9.116	187992	20.064 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.964	466405	27.276 ng/ml
2) t 2,2,4-Trimethylpentane	7.730	464430	27.487 ng/ml
3) t n-Heptane	8.070	150276	9.328 ng/ml
4) t Benzene	8.210	187777	9.266 ng/ml
6) t Toluene	10.965	466255	28.286 ng/ml
7) t Ethylbenzene	13.415	149163	8.941 ng/ml
8) t m-Xylene	13.547	315262	16.817 ng/ml
9) t O-Xylene	14.279	327791	16.128 ng/ml
10) t 1,2,4-Trimethylbenzene	16.562	401306	17.919 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB081418\
Data File : FB017390.D
Signal(s) : FID2B.CH
Acq On : 15 Aug 2018 13:44
Operator : MJ/SJ
Sample : 20 PPB GRO STD
Misc :
ALS Vial : 43 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Aug 15 14:07:45 2018
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB080818.M
Quant Title :
QLast Update : Wed Aug 08 11:39:03 2018
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#1 2-Methylpentane

R.T.: 4.964 min
Delta R.T.: 0.000 min
Response: 466405
Conc: 27.28 ng/ml




#2 2,2,4-Trimethylpentane

R.T.: 7.730 min
Delta R.T.: 0.001 min
Response: 464430
Conc: 27.49 ng/ml




#3 n-Heptane

R.T.: 8.070 min
Delta R.T.: 0.002 min
Response: 150276
Conc: 9.33 ng/ml




#4 Benzene

R.T.: 8.210 min
Delta R.T.: 0.002 min
Response: 187777
Conc: 9.27 ng/ml




#5 AAA-TFT

R.T.: 9.116 min
Delta R.T.: 0.003 min
Response: 187992
Conc: 20.06 ng/ml




#6 Toluene

R.T.: 10.965 min
Delta R.T.: 0.003 min
Response: 466255
Conc: 28.29 ng/ml




#7 Ethylbenzene

R.T.: 13.415 min
Delta R.T.: 0.003 min
Response: 149163
Conc: 8.94 ng/ml




#8 m-Xylene

R.T.: 13.547 min
Delta R.T.: 0.003 min
Response: 315262
Conc: 16.82 ng/ml



#9 O-Xylene

R.T.: 14.279 min
Delta R.T.: 0.002 min
Response: 327791
Conc: 16.13 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.562 min
Delta R.T.: 0.003 min
Response: 401306
Conc: 17.92 ng/ml