

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB090325\
Data File : FB032192.D
Signal(s) : FID2B.CH
Acq On : 3 Sep 2025 11:21
Operator : YP/AJ
Sample : VBF0903S1
Misc : 5.00G/5.00ML DI WATER
ALS Vial : 2 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Sep 04 05:13:12 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB082125.M
Quant Title :
QLast Update : Thu Aug 21 10:58:34 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.798	315450	18.330 ng/ml
Target Compounds			
6) t Toluene	10.623	5325	0.162 ng/ml
9) t O-Xylene	13.910	3161	0.111 ng/ml
10) t 1,2,4-Trimethylbenzene	16.199	10908	0.420 ng/ml


(f)=RT Delta > 1/2 Window


(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB090325\
Data File : FB032192.D
Signal(s) : FID2B.CH
Acq On : 3 Sep 2025 11:21
Operator : YP/AJ
Sample : VBF0903S1
Misc : 5.00G/5.00ML DI WATER
ALS Vial : 2 Sample Multiplier: 1

Integration File: Calibration.e
Quant Time: Sep 04 05:13:12 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB082125.M
Quant Title :
QLast Update : Thu Aug 21 10:58:34 2025
Response via : Initial Calibration
Integrator: ChemStation


Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#5 AAA-TFT

R.T.: 8.798 min
Delta R.T.: 0.001 min
Response: 315450
Conc: 18.33 ng/ml




#6 Toluene

R.T.: 10.623 min
Delta R.T.: -0.002 min
Response: 5325
Conc: 0.16 ng/ml



#9 O-Xylene

R.T.: 13.910 min
Delta R.T.: -0.010 min
Response: 3161
Conc: 0.11 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.199 min
Delta R.T.: 0.003 min
Response: 10908
Conc: 0.42 ng/ml