

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB100825\
Data File : FB032242.D
Signal(s) : FID2B.CH
Acq On : 8 Oct 2025 12:47
Operator : YP/AJ
Sample : Q3305-01
Misc : 5.00G/5.00 ML DI WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
AU-06-100725

Integration File: Calibrat.e
Quant Time: Oct 09 04:10:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB092225.M
Quant Title :
QLast Update : Mon Sep 22 12:16:10 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.804	1683982	94.080 ng/ml
Target Compounds			

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB100825\
Data File : FB032242.D
Signal(s) : FID2B.CH
Acq On : 8 Oct 2025 12:47
Operator : YP/AJ
Sample : Q3305-01
Misc : 5.00G/5.00 ML DI WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
AU-06-100725

Integration File: Calibrat.e
Quant Time: Oct 09 04:10:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB092225.M
Quant Title :
QLast Update : Mon Sep 22 12:16:10 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um





#5 AAA-TFT

R.T.: 8.804 min
Delta R.T.: 0.008 min
Response: 1683982
Conc: 94.08 ng/ml

Instrument :

FID_B

ClientSampleId :

AU-06-100725