

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB101725\
Data File : FB032254.D
Signal(s) : FID2B.CH
Acq On : 17 Oct 2025 10:05
Operator : YP/AJ
Sample : VBF1017S1
Misc : 5.00G/5.00 ML DI WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :

Integration File: Calibrat.e
Quant Time: Oct 17 23:03:24 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB092225.M
Quant Title :
QLast Update : Mon Sep 22 12:16:10 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.791	367837	20.550 ng/ml
Target Compounds			
9) t O-Xylene	13.909	3856	0.118 ng/ml
10) t 1,2,4-Trimethylbenzene	16.199	13561	0.485 ng/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB101725\
Data File : FB032254.D
Signal(s) : FID2B.CH
Acq On : 17 Oct 2025 10:05
Operator : YP/AJ
Sample : VBF1017S1
Misc : 5.00G/5.00 ML DI WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :

Integration File: Calibrat.e
Quant Time: Oct 17 23:03:24 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB092225.M
Quant Title :
QLast Update : Mon Sep 22 12:16:10 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um






#5 AAA-TFT


R.T.: 8.791 min
Delta R.T.: -0.005 min
Response: 367837
Conc: 20.55 ng/ml

Instrument :
FID_B
ClientSampleId :



#9 O-Xylene

R.T.: 13.909 min
Delta R.T.: -0.014 min
Response: 3856
Conc: 0.12 ng/ml



#10 1,2,4-Trimethylbenzene

R.T.: 16.199 min
Delta R.T.: 0.002 min
Response: 13561
Conc: 0.48 ng/ml