

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB101725\
Data File : FB032258.D
Signal(s) : FID2B.CH
Acq On : 17 Oct 2025 12:12
Operator : YP/AJ
Sample : Q3367-01
Misc : 5.00G/5.00 ML DI WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
BU-03-101625

Integration File: Calibrat.e
Quant Time: Oct 17 23:04:10 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB092225.M
Quant Title :
QLast Update : Mon Sep 22 12:16:10 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.795	328416	18.348 ng/ml
Target Compounds			

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB101725\
Data File : FB032258.D
Signal(s) : FID2B.CH
Acq On : 17 Oct 2025 12:12
Operator : YP/AJ
Sample : Q3367-01
Misc : 5.00G/5.00 ML DI WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
BU-03-101625

Integration File: Calibrat.e
Quant Time: Oct 17 23:04:10 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB092225.M
Quant Title :
QLast Update : Mon Sep 22 12:16:10 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um





#5 AAA-TFT

R.T.: 8.795 min
Delta R.T.: 0.000 min
Response: 328416
Conc: 18.35 ng/ml

Instrument :

FID_B

ClientSampleId :

BU-03-101625