

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB111425\
 Data File : FB032382.D
 Signal(s) : FID2B.CH
 Acq On : 14 Nov 2025 11:36
 Operator : YP/AJ
 Sample : BSF1114S2
 Misc : 5.00G/5.00 ML DI WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 FID_B
 ClientSampleId :
 BSF1114S2

Integration File: Calibrat.e
 Quant Time: Nov 14 14:28:27 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB102725.M
 Quant Title :
 QLast Update : Tue Oct 28 04:57:39 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 5 g/ml
 Signal Phase : RTX-502.2
 Signal Info : 60mx0.53mmx3.00um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
5) s AAA-TFT	8.794	360681	17.783 ng/ml
Target Compounds			
1) t 2-Methylpentane	4.712	670329	24.469 ng/ml
2) t 2,2,4-Trimethylpentane	7.421	883438	28.674 ng/ml
3) t n-Heptane	7.754	262489	8.958 ng/ml
4) t Benzene	7.893	349080	10.084 ng/ml
6) t Toluene	10.623	1016825	28.418 ng/ml
7) t Ethylbenzene	13.059	305860	9.617 ng/ml
8) t m-Xylene	13.192	639643	19.207 ng/ml
9) t O-Xylene	13.919	617614	19.233 ng/ml
10) t 1,2,4-Trimethylbenzene	16.194	534200	19.306 ng/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_B\Data\FB111425\
Data File : FB032382.D
Signal(s) : FID2B.CH
Acq On : 14 Nov 2025 11:36
Operator : YP/AJ
Sample : BSF1114S2
Misc : 5.00G/5.00 ML DI WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
FID_B
ClientSampleId :
BSF1114S2

Integration File: Calibrat.e
Quant Time: Nov 14 14:28:27 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_B\Method\FB102725.M
Quant Title :
QLast Update : Tue Oct 28 04:57:39 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 5 g/ml
Signal Phase : RTX-502.2
Signal Info : 60mx0.53mmx3.00um

