

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC010323AL\
 Data File : FC062551.D
 Signal(s) : FID1A.ch
 Acq On : 03 Jan 2023 19:39
 Operator : YP/AJ
 Sample : N6248-12D
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :
 TP-20

Integration File: autoint1.e
 Quant Time: Jan 04 02:45:45 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 120722.M
 Quant Title : GC Extractables
 QLast Update : Thu Dec 08 01:05:24 2022
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.560	4528095	32.100 ug/ml
Spiked Amount 50.000		Recovery =	64.20%
12) S 1-chlorooctadecane (S...	12.998	3429277	32.174 ug/ml
Spiked Amount 50.000		Recovery =	64.35%

Target Compounds

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC010323AL\
Data File : FC062551.D
Signal(s) : FID1A.ch
Acq On : 03 Jan 2023 19:39
Operator : YP/AJ
Sample : N6248-12D
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :
TP-20

Integration File: autoint1.e
Quant Time: Jan 04 02:45:45 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 120722.M
Quant Title : GC Extractables
QLast Update : Thu Dec 08 01:05:24 2022
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.560 min
Delta R.T.: -0.012 min
Response: 4528095
Conc: 32.10 ug/ml

Instrument :
FID_C
ClientSampleId :
TP-20



#12 1-chlorooctadecane (SURR)

R.T.: 12.998 min
Delta R.T.: -0.014 min
Response: 3429277
Conc: 32.17 ug/ml