

Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_C\Data\FC010521AL\  
 Data File : FC051558.D  
 Signal(s) : FID1A.CH  
 Acq On : 05 Jan 2021 10:27  
 Operator : DD\AJ  
 Sample : L5249-01RE  
 Misc :  
 ALS Vial : 3 Sample Multiplier: 1

Instrument :  
 FID\_C  
 ClientSampleId :  
 BU-03-123120RE

Integration File: autoint1.e  
 Quant Time: Jan 06 07:59:38 2021  
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_C\Method\Alaphatic EPH 122120.M  
 Quant Title : GC Extractables  
 QLast Update : Tue Dec 22 00:33:00 2020  
 Response via : Initial Calibration  
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul  
 Signal Phase : Rxi-1ms  
 Signal Info : 20M x 0.18mm x 0.18um

| Compound                       | R.T.   | Response   | Conc Units    |
|--------------------------------|--------|------------|---------------|
| -----                          |        |            |               |
| System Monitoring Compounds    |        |            |               |
| 9) S ortho-Terphenyl (SURR)    | 11.563 | 4990502    | 43.059 ug/ml  |
| Spiked Amount 50.000           |        | Recovery = | 86.12%        |
| 12) S 1-chlorooctadecane (S... | 12.991 | 15695171   | 173.985 ug/ml |
| Spiked Amount 50.000           |        | Recovery = | 347.97%       |

Target Compounds

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID\_C\Data\FC010521AL\  
Data File : FC051558.D  
Signal(s) : FID1A.CH  
Acq On : 05 Jan 2021 10:27  
Operator : DD\AJ  
Sample : L5249-01RE  
Misc :  
ALS Vial : 3 Sample Multiplier: 1

Instrument :  
FID\_C  
ClientSampleId :  
BU-03-123120RE

Integration File: autoint1.e  
Quant Time: Jan 06 07:59:38 2021  
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID\_C\Method\Alaphatic EPH 122120.M  
Quant Title : GC Extractables  
QLast Update : Tue Dec 22 00:33:00 2020  
Response via : Initial Calibration  
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul  
Signal Phase : Rxi-1ms  
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.563 min  
Delta R.T.: -0.006 min  
Response: 4990502  
Conc: 43.06 ug/ml

Instrument :  
FID\_C  
ClientSampleId :  
BU-03-123120RE



#12 1-chlorooctadecane (SURR)

R.T.: 12.991 min  
Delta R.T.: -0.014 min  
Response: 15695171  
Conc: 173.99 ug/ml