

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC011322AL\
 Data File : FC057903.D
 Signal(s) : FID1A.CH
 Acq On : 14 Jan 2022 5:55
 Operator : AJ\MA
 Sample : N1103-01DL2 50X
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :
 CL-01-011122DL2

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 01/14/2022
 Supervised By :Ankita Jodhani 01/14/2022

Integration File: sample.E
 Quant Time: Jan 14 10:11:21 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 010522.M
 Quant Title : GC Extractables
 QLast Update : Wed Jan 05 12:13:43 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	10.705	120163	0.621 ug/ml
Spiked Amount 50.000		Recovery =	1.24%
12) S 1-chlorooctadecane (S...	12.144	81228	0.551 ug/mlm
Spiked Amount 50.000		Recovery =	1.10%

Target Compounds

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC011322AL\
Data File : FC057903.D
Signal(s) : FID1A.CH
Acq On : 14 Jan 2022 5:55
Operator : AJ\MA
Sample : N1103-01DL2 50X
Misc :
ALS Vial : 34 Sample Multiplier: 1


Instrument :
FID_C
ClientSampleId :
CL-01-011122DL2


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 01/14/2022
Supervised By :Ankita Jodhani 01/14/2022

Integration File: sample.E
Quant Time: Jan 14 10:11:21 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 010522.M
Quant Title : GC Extractables
QLast Update : Wed Jan 05 12:13:43 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 10.705 min
Delta R.T.: -0.013 min
Response: 120163
Conc: 0.62 ug/ml

Instrument :
FID_C
ClientSampleId :
CL-01-011122DL2

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 01/14/2022
Supervised By :Ankita Jodhani 01/14/2022



#12 1-chlorooctadecane (SURR)

R.T.: 12.144 min
Delta R.T.: -0.011 min
Response: 81228
Conc: 0.55 ug/ml m