

Data Path : P:\HPCHEM1\FID_C\Data\FC011618AL\
 Data File : FC031776.D
 Signal(s) : FID1A.CH
 Acq On : 16 Jan 2018 15:05
 Operator : UA/AJ
 Sample : 20 PPM ALIPHATIC HC STD
 Misc :
 ALS Vial : 52 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Jan 16 15:45:57 2018
 Quant Method : P:\HPCHEM1\FID_C\Method\Alaphatic EPH 011618.M
 Quant Title : GC Extractables
 QLast Update : Tue Jan 16 13:50:30 2018
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 T	n-Nonane (C9)	123.250	129.970 E3	-5.5	114	0.00
2 T	n-Decane (C10)	121.576	127.861 E3	-5.2	113	0.00
3 T	A-Naphthalene (C11.7)	128.946	135.215 E3	-4.9	112	0.00
4 T	n-Dodecane (C12)	121.538	126.978 E3	-4.5	113	0.00
5 T	A-2-methylnaphthalene (C12.7)	125.571	133.306 E3	-6.2	113	0.00
6 T	n-Tetradecane (C14)	121.313	125.879 E3	-3.8	112	0.00
7 T	n-Hexadecane (C16)	120.201	123.611 E3	-2.8	111	0.00
8 T	n-Octadecane (C18)	117.659	120.707 E3	-2.6	111	0.00
9 S	ortho-Terphenyl (SURR)	128.462	132.125 E3	-2.9	110	0.00
10 T	n-Eicosane (C20)	118.037	121.244 E3	-2.7	110	0.00
11 T	n-Heneicosane (C21)	116.977	120.158 E3	-2.7	110	0.00
12 S	1-chlorooctadecane (SURR)	93.409	95.854 E3	-2.6	110	0.00
13 T	n-Docosane (C22)	116.787	120.352 E3	-3.1	111	0.00
14 T	n-Tetracosane (C24)	114.581	119.715 E3	-4.5	112	0.00
15 T	n-Hexacosane (C26)	113.929	118.887 E3	-4.4	112	0.00
16 T	n-Octacosane (C28)	115.072	119.767 E3	-4.1	112	0.00
17 T	n-Tricontane (C30)	111.601	115.578 E3	-3.6	112	0.00
18 T	n-Dotriacontane (C32)	109.584	112.224 E3	-2.4	112	0.00
19 T	n-Tetratriacontane (C34)	103.401	105.906 E3	-2.4	111	0.00
20 T	n-Hexatriacontane (C36)	96.431	97.589 E3	-1.2	107	0.00
21 T	n-Octatriacontane (C38)	92.053	93.781 E3	-1.9	109	0.00
22 T	n-Tetracontane (C40)	88.629	87.904 E3	0.8	103	0.00

Evaluate Continuing Calibration Report - Not Found

(#) = Out of Range

SPCC's out = 0 CCC's out = 0