

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC011723AL\
 Data File : FC062684.D
 Signal(s) : FID1A.ch
 Acq On : 17 Jan 2023 09:09
 Operator : YP/AJ
 Sample : 01161-01DL 5X
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :

Integration File: autoint1.e
 Quant Time: Jan 17 11:02:23 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 010523.M
 Quant Title : GC Extractables
 QLast Update : Thu Jan 05 04:54:00 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.552	61044	0.432 ug/ml
Spiked Amount 50.000		Recovery =	0.86%
12) S 1-chlorooctadecane (S...	12.992	481358	4.570 ug/ml
Spiked Amount 50.000		Recovery =	9.14%
Target Compounds			
1) T n-Nonane (C9)	3.311	2346041	17.544 ug/ml
2) T n-Decane (C10)	4.378	1787225	13.780 ug/ml
3) T A~Naphthalene (C11.7)	5.964	212244	1.444 ug/ml
4) T n-Dodecane (C12)	6.399	610974	4.754 ug/ml
5) T A~2-methylnaphthalene...	7.003	29928	0.219 ug/ml
6) T n-Tetradecane (C14)	8.196	101318	0.787 ug/ml
10) T n-Eicosane (C20)	12.548	39355	0.319 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC011723AL\
Data File : FC062684.D
Signal(s) : FID1A.ch
Acq On : 17 Jan 2023 09:09
Operator : YP/AJ
Sample : 01161-01DL 5X
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :

Integration File: autoint1.e
Quant Time: Jan 17 11:02:23 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 010523.M
Quant Title : GC Extractables
QLast Update : Thu Jan 05 04:54:00 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#1 n-Nonane (C9)


R.T.: 3.311 min
Delta R.T.: -0.005 min
Response: 2346041
Conc: 17.54 ug/ml

Instrument :
FID_C
ClientSampleId :




#2 n-Decane (C10)

R.T.: 4.378 min
Delta R.T.: -0.005 min
Response: 1787225
Conc: 13.78 ug/ml




#3 A~Naphthalene (C11.7)

R.T.: 5.964 min
Delta R.T.: -0.004 min
Response: 212244
Conc: 1.44 ug/ml



#4 n-Dodecane (C12)


R.T.: 6.399 min
Delta R.T.: -0.004 min
Response: 610974
Conc: 4.75 ug/ml



#5 A~2-methylnaphthalene (C12.89)


R.T.: 7.003 min
Delta R.T.: -0.023 min
Response: 29928
Conc: 0.22 ug/ml

Instrument :
FID_C
ClientSampleId :




#6 n-Tetradecane (C14)

R.T.: 8.196 min
Delta R.T.: -0.004 min
Response: 101318
Conc: 0.79 ug/ml




#9 ortho-Terphenyl (SURR)

R.T.: 11.552 min
Delta R.T.: -0.007 min
Response: 61044
Conc: 0.43 ug/ml



#10 n-Eicosane (C20)

R.T.: 12.548 min
Delta R.T.: -0.005 min
Response: 39355
Conc: 0.32 ug/ml



#12 1-chlorooctadecane (SURR)

R.T.: 12.992 min
Delta R.T.: -0.005 min
Response: 481358
Conc: 4.57 ug/ml

Instrument :
FID_C
ClientSampleId :