

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC013123AL\
 Data File : FC062815.D
 Signal(s) : FID1A.ch
 Acq On : 31 Jan 2023 12:37
 Operator : YP/AJ
 Sample : 01329-13
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :

Integration File: autoint1.e
 Quant Time: Jan 31 13:53:13 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 012623.M
 Quant Title : GC Extractables
 QLast Update : Fri Jan 27 04:11:29 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.537	360025	2.895 ug/ml
Spiked Amount 50.000		Recovery =	5.79%
12) S 1-chlorooctadecane (S...	12.978	2402642	25.474 ug/ml
Spiked Amount 50.000		Recovery =	50.95%

Target Compounds

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC013123AL\
Data File : FC062815.D
Signal(s) : FID1A.ch
Acq On : 31 Jan 2023 12:37
Operator : YP/AJ
Sample : 01329-13
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :

Integration File: autoint1.e
Quant Time: Jan 31 13:53:13 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 012623.M
Quant Title : GC Extractables
QLast Update : Fri Jan 27 04:11:29 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.537 min
Delta R.T.: -0.005 min
Response: 360025
Conc: 2.90 ug/ml

Instrument :
FID_C
ClientSampleId :



#12 1-chlorooctadecane (SURR)

R.T.: 12.978 min
Delta R.T.: -0.004 min
Response: 2402642
Conc: 25.47 ug/ml