

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC013123AL\
 Data File : FC062828.D
 Signal(s) : FID1A.ch
 Acq On : 31 Jan 2023 22:05
 Operator : YP/AJ
 Sample : 01352-01
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :

Integration File: autoint1.e
 Quant Time: Feb 01 03:20:53 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 012623.M
 Quant Title : GC Extractables
 QLast Update : Fri Jan 27 04:11:29 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.538	3888538	31.272 ug/ml
Spiked Amount 50.000		Recovery =	62.54%
12) S 1-chlorooctadecane (S...	12.977	3033836	32.166 ug/ml
Spiked Amount 50.000		Recovery =	64.33%

Target Compounds

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC013123AL\
Data File : FC062828.D
Signal(s) : FID1A.ch
Acq On : 31 Jan 2023 22:05
Operator : YP/AJ
Sample : 01352-01
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :

Integration File: autoint1.e
Quant Time: Feb 01 03:20:53 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 012623.M
Quant Title : GC Extractables
QLast Update : Fri Jan 27 04:11:29 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.538 min
Delta R.T.: -0.004 min
Response: 3888538
Conc: 31.27 ug/ml

Instrument :
FID_C
ClientSampleId :



#12 1-chlorooctadecane (SURR)

R.T.: 12.977 min
Delta R.T.: -0.004 min
Response: 3033836
Conc: 32.17 ug/ml