

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC020723AL\
 Data File : FC062889.D
 Signal(s) : FID1A.ch
 Acq On : 07 Feb 2023 23:04
 Operator : YP/AJ
 Sample : 01463-15
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :

Integration File: autoint1.e
 Quant Time: Feb 08 01:16:21 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 012623.M
 Quant Title : GC Extractables
 QLast Update : Fri Jan 27 04:11:29 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.544	2043398	16.433 ug/ml
Spiked Amount	50.000	Recovery	= 32.87%
12) S 1-chlorooctadecane (S...	12.986	5318527	56.389 ug/ml
Spiked Amount	50.000	Recovery	= 112.78%

Target Compounds

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC020723AL\
Data File : FC062889.D
Signal(s) : FID1A.ch
Acq On : 07 Feb 2023 23:04
Operator : YP/AJ
Sample : 01463-15
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :

Integration File: autoint1.e
Quant Time: Feb 08 01:16:21 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 012623.M
Quant Title : GC Extractables
QLast Update : Fri Jan 27 04:11:29 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.544 min
Delta R.T.: 0.002 min
Response: 2043398
Conc: 16.43 ug/ml

Instrument :
FID_C
ClientSampleId :



#12 1-chlorooctadecane (SURR)

R.T.: 12.986 min
Delta R.T.: 0.005 min
Response: 5318527
Conc: 56.39 ug/ml