

Sample Name: S0 Acquired: 02/07/2019 11:24:34 Type: Cal
Method: ISM02.4 INSTRUMENT P5 WITH 6 POINT ICAL(v774) Mode: IR Corr. Factor: 1.000000
User: BIN H Custom ID1: S00 Custom ID2: Custom ID3:
Comment:

Elem	As1890	Tl1908	Pb2203	Se1960	Sb2068	Al3961	Ba4934	Be2348
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	-.00028	-.00066	-.00018	.00055	.00041	.00443	.02223	-.00329
Stddev	.00013	.00006	.00022	.00008	.00024	.00041	.00458	.00047
%RSD	46.765	8.4180	118.34	15.102	57.526	9.3263	20.594	14.328
#1	-.00018	-.00059	-.00011	.00064	.00016	.00430	.01747	-.00275
#2	-.00042	-.00069	-.00001	.00051	.00063	.00489	.02660	-.00351
#3	-.00022	-.00069	-.00042	.00049	.00043	.00409	.02264	-.00361
Elem	Cd2144	Ca3736	Cr2677	Co2286	Cu3247	Fe2598	Mn2576	Mg2790
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	.00051	.00346	.00004	.00036	.00117	.00036	-.00009	-.00011
Stddev	.00021	.00043	.00009	.00014	.00049	.00012	.00010	.00015
%RSD	40.348	12.438	193.81	39.182	42.032	34.702	108.06	146.85
#1	.00031	.00348	-.00005	.00048	.00062	.00034	-.00021	-.00021
#2	.00051	.00388	.00012	.00021	.00134	.00024	-.00005	.00007
#3	.00072	.00302	.00006	.00038	.00156	.00049	-.00002	-.00017
Elem	Ni2316	Ag3280	Na8183	V_2924	Zn2138	K_7698	P_1774	B_2496
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	-.00199	-.00125	.00236	-.00029	.00018	.02800	-.00106	.00578
Stddev	.00002	.00007	.00131	.00034	.00036	.00222	.00004	.00053
%RSD	1.2274	5.9895	55.381	115.67	204.31	7.9165	3.7973	9.2129
#1	-.00202	-.00133	.00332	.00000	-.00001	.02558	-.00102	.00609
#2	-.00197	-.00120	.00290	-.00066	.00059	.02993	-.00110	.00517
#3	-.00199	-.00121	.00087	-.00021	-.00005	.02851	-.00107	.00610
Elem	Mo2020	S_1820	Si2516	Sn1899	Ti3361	Li6707	Sr4077	
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S	
Avg	.00077	.00029	.01200	.00034	.00185	.01811	.00101	
Stddev	.00007	.00009	.00094	.00017	.00068	.00291	.00077	
%RSD	9.3931	32.581	7.8069	49.740	36.608	16.090	75.692	
#1	.00081	.00039	.01112	.00051	.00259	.01924	.00116	
#2	.00080	.00021	.01190	.00035	.00127	.02030	.00169	
#3	.00068	.00026	.01299	.00017	.00167	.01480	.00018	

Sample Name: S0 Acquired: 02/07/2019 11:24:34 Type: Cal
Method: ISM02.4 INSTRUMENT P5 WITH 6 POINT ICAL(v774) Mode: IR Corr. Factor: 1.000000
User: BIN H Custom ID1: S00 Custom ID2: Custom ID3:
Comment:

Int. Std.	Y_2243	Y_2243	Y_3600	Y_3710	In2306
Units	Cts/S	Cts/S	Cts/S	Cts/S	Cts/S
Avg	1886.9	2155.7	38954.	3614.0	3178.6
Stddev	18.4	20.6	262.	5.7	31.6
%RSD	.97327	.95434	.67202	.15703	.99419
#1	1866.6	2173.6	38756.	3608.5	3208.0
#2	1891.7	2160.4	38855.	3619.8	3182.7
#3	1902.3	2133.3	39251.	3613.8	3145.2