

Instrument :
FID_C
LabSampleId :
20 PPM ALIPHATIC HC STD

rteres

Area Percent Report

Manual IntegrationsAPPROVED

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC02232
Data File : FC058432.D
Signal(s) : FID1A.CH
Acq On : 23 Feb 2022 20:55
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 4 Sample Multiplier: 1

Reviewed By :Yogesh Patel 02/24/2022
Supervised By :Ankita Jodhani 02/24/2022

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 022322.M
Title : GC Extractables

Signal : FID1A.CH

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.246	3.211	3.331	BB	280111	2403907	82.49%	4.600%
2	4.295	4.261	4.385	BB	262068	2382838	81.77%	4.560%
3	5.854	5.828	6.028	BB	244451	2650249	90.94%	5.071%
4	6.289	6.248	6.375	BB	246429	2465116	84.59%	4.717%
5	6.899	6.865	7.061	BB	242464	2651770	91.00%	5.074%
6	8.067	8.035	8.168	BB	238190	2513897	86.27%	4.810%
7	9.654	9.601	9.765	BB	233606	2583292	88.65%	4.943%
8	11.086	11.041	11.175	BB	227903	2638778	90.55%	5.049%
9	11.391	11.355	11.478	BB	271764	2914152	100.00%	5.576%
10	12.385	12.348	12.465	BB	220670	2622829	90.00%	5.019%
11	12.825	12.781	12.918	BB	187269	2272267	77.97%	4.348%
12	12.994	12.955	13.088	BB	213425	2634088	90.39%	5.040%
13	13.576	13.538	13.648	BB	205505	2599460	89.20%	4.974%
14	14.671	14.615	14.748	BB	197590	2552953	87.61%	4.885%
15	15.686	15.645	15.761	BB	188877	2463473	84.53%	4.714%
16	16.631	16.508	16.695	BB	173695	2362329	81.06%	4.520%
17	17.513	17.458	17.585	BB	150459	2257940	77.48%	4.321%
18	18.342	18.165	18.418	BB	146912	2120950	72.78%	4.059%
19	19.123	19.081	19.178	BB	131775	1959467	67.24%	3.750%
20	19.859	19.821	19.895	BV	126660	1811639	62.17%	3.467%
21	20.563	20.528	20.606	BB	107593	1724861	59.19%	3.301%
22	21.388	21.345	21.438	BV	80312	1673129	57.41%	3.202%
Sum of corrected areas:						52259384		

Aliphatic EPH 022322.M Thu Feb 24 08:50:45 2022