

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC022420AL\
 Data File : FC045933.D
 Signal(s) : FID1A.CH
 Acq On : 24 Feb 2020 22:16
 Operator : DD\AJ
 Sample : 20 PPM ALIPHATIC HC STD
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 FID_C
 LabSampleId :
 20 PPM ALIPHATIC HC STD

Integration File: sample.E
 Quant Time: Feb 25 00:43:25 2020
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Alaphatic EPH 022120.M
 Quant Title : GC Extractables
 QLast Update : Sat Feb 22 03:16:45 2020
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 T n-Nonane (C9)	100.802	111.074 E3	-10.2	105	0.00
2 T n-Decane (C10)	100.862	112.217 E3	-11.3	106	0.00
3 T A~Naphthalene (C11.7)	111.036	123.361 E3	-11.1	105	0.00
4 T n-Dodecane (C12)	102.020	113.953 E3	-11.7	106	0.00
5 T A~2-methylnaphthalene (C12.	111.236	124.816 E3	-12.2	106	0.00
6 T n-Tetradecane (C14)	102.368	114.079 E3	-11.4	106	0.00
7 T n-Hexadecane (C16)	103.537	114.832 E3	-10.9	106	0.00
8 T n-Octadecane (C18)	105.123	115.869 E3	-10.2	105	0.00
9 S ortho-Terphenyl (SURR)	114.079	125.800 E3	-10.3	105	0.00
10 T n-Eicosane (C20)	105.710	116.075 E3	-9.8	105	0.00
11 T n-Heneicosane (C21)	106.659	116.870 E3	-9.6	105	0.00
12 S 1-chlorooctadecane (SURR)	89.710	98.307 E3	-9.6	105	0.00
13 T n-Docosane (C22)	106.257	116.326 E3	-9.5	105	0.00
14 T n-Tetracosane (C24)	107.027	116.739 E3	-9.1	104	0.00
15 T n-Hexacosane (C26)	106.284	115.410 E3	-8.6	104	0.00
16 T n-Octacosane (C28)	104.697	113.108 E3	-8.0	103	0.00
17 T n-Tricontane (C30)	100.772	108.382 E3	-7.6	103	0.00
18 T n-Dotriacontane (C32)	85.584	91.416 E3	-6.8	101	0.00
19 T n-Tetratriacontane (C34)	72.579	77.204 E3	-6.4	102	0.00
20 T n-Hexatriacontane (C36)	67.628	71.074 E3	-5.1	100	0.00
21 T n-Octatriacontane (C38)	60.442	63.669 E3	-5.3	100	0.00
22 T n-Tetracontane (C40)	62.871	66.675 E3	-6.1	102	0.00

Evaluate Continuing Calibration Report - Not Found

(#) = Out of Range

SPCC's out = 0 CCC's out = 0