

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC022619AL\
 Data File : FC039426.D
 Signal(s) : FID1A.CH
 Acq On : 26 Feb 2019 23:28
 Operator : AJ\SJ
 Sample : 20 PPM ALIPHATIC HC STD
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Integration File: sample.E
 Quant Time: Feb 27 02:23:30 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Alaphatic EPH 022219.M
 Quant Title : GC Extractables
 QLast Update : Mon Feb 25 18:21:16 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound		AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 T	n-Nonane (C9)	120.021	111.015 E3	7.5	96	0.00
2 T	n-Decane (C10)	120.473	111.433 E3	7.5	95	0.00
3 T	A~Naphthalene (C11.7)	130.125	118.347 E3	9.1	93	-0.01
4 T	n-Dodecane (C12)	123.548	112.035 E3	9.3	93	0.00
5 T	A~2-methylnaphthalene (C12.	129.684	110.488 E3	14.8	87	-0.01
6 T	n-Tetradecane (C14)	127.365	111.210 E3	12.7	88	0.00
7 T	n-Hexadecane (C16)	129.589	111.297 E3	14.1	86	0.00
8 T	n-Octadecane (C18)	134.523	111.428 E3	17.2	83	0.00
9 S	ortho-Terphenyl (SURR)	168.201	147.731 E3	12.2	87	0.00
10 T	n-Eicosane (C20)	130.397	112.175 E3	14.0	85	0.00
11 T	n-Heneicosane (C21)	126.822	112.809 E3	11.0	88	0.00
12 S	1-chlorooctadecane (SURR)	126.142	109.807 E3	12.9	86	0.00
13 T	n-Docosane (C22)	125.967	111.475 E3	11.5	88	0.00
14 T	n-Tetracosane (C24)	124.986	111.414 E3	10.9	90	0.00
15 T	n-Hexacosane (C26)	124.735	111.232 E3	10.8	93	0.00
16 T	n-Octacosane (C28)	133.655	112.680 E3	15.7	91	0.00
17 T	n-Tricontane (C30)	146.758	112.843 E3	23.1	86	0.00
18 T	n-Dotriacontane (C32)	134.335	110.228 E3	17.9	91	0.00
19 T	n-Tetratriacontane (C34)	113.746	97.643 E3	14.2	94	-0.01
20 T	n-Hexatriacontane (C36)	96.447	84.895 E3	12.0	95	-0.01
21 T	n-Octatriacontane (C38)	91.816	80.005 E3	12.9	96	-0.01
22 T	n-Tetracontane (C40)	90.758	77.778 E3	14.3	93	-0.01

Evaluate Continuing Calibration Report - Not Found

(#) = Out of Range

SPCC's out = 0 CCC's out = 0