

Instrument :
FID_C
LabSampleId :
20 PPM ALIPHATIC HC STD

nteres

Area Percent Report

Manual IntegrationsAPPROVED

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\F0
Data File : FC065870.D
Signal(s) : FID1A.ch
Acq On : 26 Feb 2024 23:20
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 2 Sample Multiplier: 1

Reviewed By :Yogesh Patel 02/27/2024
Supervised By :Ankita Jodhani 02/27/2024

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 020824.M
Title : GC Extractables

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.286	3.261	3.398	BV	321882	2829729	83.06%	4.269%
2	4.348	4.301	4.511	BB	306744	2966997	87.09%	4.476%
3	5.936	5.841	6.298	BB	199440	3134864	92.02%	4.729%
4	6.362	6.328	6.561	BB	292194	3017203	88.57%	4.551%
5	6.993	6.966	7.254	PV	174493	2914885	85.56%	4.397%
6	8.152	7.951	8.314	BB	264096	2931037	86.04%	4.421%
7	9.746	9.710	9.938	PB	251882	3041370	89.28%	4.588%
8	11.182	11.144	11.394	BB	251589	3089226	90.68%	4.660%
9	11.496	11.431	11.644	BV	292758	3327110	97.66%	5.019%
10	12.484	12.371	12.591	BV	248451	2965713	87.06%	4.474%
11	12.927	12.808	13.061	BV	200118	2515688	73.85%	3.795%
12	13.093	13.061	13.218	VB	238788	2957862	86.83%	4.462%
13	13.675	13.611	13.811	BB	238050	2967883	87.12%	4.477%
14	14.774	14.711	14.918	BB	218369	2988920	87.74%	4.509%
15	15.788	15.691	15.898	BB	216937	2981422	87.52%	4.497%
16	16.732	16.661	16.848	BB	225125	3071475	90.16%	4.633%
17	17.616	17.548	17.721	VB	255992	3349015	98.31%	5.052%
18	18.443	18.384	18.507	PV	237977	3406667	100.00%	5.139%
19	19.224	19.171	19.314	PB	221841	3218712	94.48%	4.855%
20	19.959	19.900	19.998	VV	204444	2998566	88.02%	4.523%
21	20.675	20.621	20.748	VV	167068	2879074	84.51%	4.343%
22	21.548	21.474	21.634	BV	112931	2739050	80.40%	4.132%
Sum of corrected areas:						66292467		

Aliphatic EPH 020824.M Tue Feb 27 05:22:07 2024