

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC030323AL\
 Data File : FC063087.D
 Signal(s) : FID1A.ch
 Acq On : 03 Mar 2023 13:42
 Operator : YP/AJ
 Sample : 01749-04
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :
 CF-6

Integration File: autoint1.e
 Quant Time: Mar 03 22:59:26 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 022223.M
 Quant Title : GC Extractables
 QLast Update : Wed Feb 22 04:36:00 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.535	6642182	51.172 ug/ml
Spiked Amount 50.000		Recovery =	102.34%
12) S 1-chlorooctadecane (S...	12.973	5207997	53.389 ug/ml
Spiked Amount 50.000		Recovery =	106.78%

Target Compounds

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC030323AL\
Data File : FC063087.D
Signal(s) : FID1A.ch
Acq On : 03 Mar 2023 13:42
Operator : YP/AJ
Sample : 01749-04
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :
CF-6

Integration File: autoint1.e
Quant Time: Mar 03 22:59:26 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 022223.M
Quant Title : GC Extractables
QLast Update : Wed Feb 22 04:36:00 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.535 min
Delta R.T.: -0.010 min
Response: 6642182
Conc: 51.17 ug/ml

Instrument :
FID_C
ClientSampleId :
CF-6



#12 1-chlorooctadecane (SURR)

R.T.: 12.973 min
Delta R.T.: -0.009 min
Response: 5207997
Conc: 53.39 ug/ml