

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC030421AL\
 Data File : FC052237.D
 Signal(s) : FID1A.CH
 Acq On : 05 Mar 2021 5:17
 Operator : DD\AJ
 Sample : M1551-01
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :

Integration File: autoint1.e
 Quant Time: Mar 05 06:30:10 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Alaphatic EPH 030321.M
 Quant Title : GC Extractables
 QLast Update : Thu Mar 04 03:56:17 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.623	5519233	41.852 ug/ml
Spiked Amount 50.000		Recovery =	83.70%
12) S 1-chlorooctadecane (S...	13.060	4392821	42.239 ug/ml
Spiked Amount 50.000		Recovery =	84.48%
Target Compounds			
10) T n-Eicosane (C20)	12.673f	336701	2.733 ug/ml
14) T n-Tetracosane (C24)	14.893	159036	1.364 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC030421AL\
Data File : FC052237.D
Signal(s) : FID1A.CH
Acq On : 05 Mar 2021 5:17
Operator : DD\AJ
Sample : M1551-01
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :

Integration File: autoint1.e
Quant Time: Mar 05 06:30:10 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Alaphatic EPH 030321.M
Quant Title : GC Extractables
QLast Update : Thu Mar 04 03:56:17 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)


R.T.: 11.623 min
Delta R.T.: 0.008 min
Response: 5519233
Conc: 41.85 ug/ml

Instrument :
FID_C
ClientSampleId :




#10 n-Eicosane (C20)

R.T.: 12.673 min
Delta R.T.: 0.063 min
Response: 336701
Conc: 2.73 ug/ml



#12 1-chlorooctadecane (SURR)

R.T.: 13.060 min
Delta R.T.: 0.006 min
Response: 4392821
Conc: 42.24 ug/ml



#14 n-Tetracosane (C24)

R.T.: 14.893 min
Delta R.T.: -0.021 min
Response: 159036
Conc: 1.36 ug/ml