

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC030519AL\
 Data File : FC039610.D
 Signal(s) : FID1A.CH
 Acq On : 05 Mar 2019 15:13
 Operator : AJ\SJ
 Sample : K1715-05
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :
 TP-4-B-D1

Integration File: autoint1.e
 Quant Time: Mar 06 02:15:02 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Alaphatic EPH 022219.M
 Quant Title : GC Extractables
 QLast Update : Mon Feb 25 18:21:16 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
12) S 1-chlorooctadecane (S...	12.605	4503658	35.703 ug/ml
Spiked Amount 50.000		Recovery =	71.41%
Target Compounds			
17) T n-Tricontane (C30)	17.295	129906	0.885 ug/ml
18) T n-Dotriacontane (C32)	18.124	100466	0.748 ug/ml
19) T n-Tetratriacontane (C34)	18.903	59441	0.523 ug/ml
20) T n-Hexatriacontane (C36)	19.639	12688	0.132 ug/ml
21) T n-Octatriacontane (C38)	20.340	411500	4.482 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC030519AL\
Data File : FC039610.D
Signal(s) : FID1A.CH
Acq On : 05 Mar 2019 15:13
Operator : AJ\SJ
Sample : K1715-05
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :
TP-4-B-D1

Integration File: autoint1.e
Quant Time: Mar 06 02:15:02 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Alaphatic EPH 022219.M
Quant Title : GC Extractables
QLast Update : Mon Feb 25 18:21:16 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)


R.T.: 0.000 min
Exp R.T.: 11.178 min
Response: 0
Conc: N.D.

Instrument :
FID_C
ClientSampleId :
TP-4-B-D1




#12 1-chlorooctadecane (SURR)

R.T.: 12.605 min
Delta R.T.: -0.016 min
Response: 4503658
Conc: 35.70 ug/ml




#17 n-Tricontane (C30)

R.T.: 17.295 min
Delta R.T.: -0.016 min
Response: 129906
Conc: 0.89 ug/ml



#18 n-Dotriacontane (C32)


R.T.: 18.124 min
Delta R.T.: -0.015 min
Response: 100466
Conc: 0.75 ug/ml



#19 n-Tetratriacontane (C34)


R.T.: 18.903 min
Delta R.T.: -0.016 min
Response: 59441
Conc: 0.52 ug/ml

Instrument :
FID_C
ClientSampleId :
TP-4-B-D1



#20 n-Hexatriacontane (C36)

R.T.: 19.639 min
Delta R.T.: -0.016 min
Response: 12688
Conc: 0.13 ug/ml



#21 n-Octatriacontane (C38)

R.T.: 20.340 min
Delta R.T.: -0.013 min
Response: 411500
Conc: 4.48 ug/ml