

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC030821AL\
 Data File : FC052288.D
 Signal(s) : FID1A.CH
 Acq On : 08 Mar 2021 8:03
 Operator : DD\AJ
 Sample : MECL2
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :

Integration File: autoint1.e
 Quant Time: Mar 09 02:54:13 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Alaphatic EPH 030321.M
 Quant Title : GC Extractables
 QLast Update : Thu Mar 04 03:56:17 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
Target Compounds			
11) T n-Heneicosane (C21)	13.214	416803	3.414 ug/ml
13) T n-Docosane (C22)	13.798	192370	1.599 ug/ml
14) T n-Tetracosane (C24)	14.904	128816	1.105 ug/ml
15) T n-Hexacosane (C26)	15.927	96707	0.880 ug/ml
16) T n-Octacosane (C28)	16.879	92049	0.898 ug/ml
18) T n-Dotriacontane (C32)	18.598	248019	2.905 ug/ml
19) T n-Tetratriacontane (C34)	19.381	118868	1.533 ug/ml
20) T n-Hexatriacontane (C36)	20.126	110660	1.540 ug/ml
21) T n-Octatriacontane (C38)	20.868	125560	1.833 ug/ml
22) T n-Tetracontane (C40)	21.813	490315	6.818 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC030821AL\
Data File : FC052288.D
Signal(s) : FID1A.CH
Acq On : 08 Mar 2021 8:03
Operator : DD\AJ
Sample : MECL2
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :

Integration File: autoint1.e
Quant Time: Mar 09 02:54:13 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Alaphatic EPH 030321.M
Quant Title : GC Extractables
QLast Update : Thu Mar 04 03:56:17 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURRE)


R.T.: 0.000 min
Exp R.T.: 11.615 min
Response: 0
Conc: N.D.

Instrument :
FID_C
ClientSampleId :




#11 n-Heneicosane (C21)

R.T.: 13.214 min
Delta R.T.: -0.008 min
Response: 416803
Conc: 3.41 ug/ml




#12 1-chlorooctadecane (SURRE)

R.T.: 0.000 min
Exp R.T.: 13.054 min
Response: 0
Conc: N.D.



#13 n-Docosane (C22)


R.T.: 13.798 min
Delta R.T.: -0.012 min
Response: 192370
Conc: 1.60 ug/ml



#14 n-Tetracosane (C24)


R.T.: 14.904 min
Delta R.T.: -0.010 min
Response: 128816
Conc: 1.10 ug/ml

Instrument :
FID_C
ClientSampleId :




#15 n-Hexacosane (C26)

R.T.: 15.927 min
Delta R.T.: -0.009 min
Response: 96707
Conc: 0.88 ug/ml




#16 n-Octacosane (C28)

R.T.: 16.879 min
Delta R.T.: -0.008 min
Response: 92049
Conc: 0.90 ug/ml



#18 n-Dotriacontane (C32)


R.T.: 18.598 min
Delta R.T.: -0.010 min
Response: 248019
Conc: 2.90 ug/ml



#19 n-Tetratriacontane (C34)


R.T.: 19.381 min
Delta R.T.: -0.011 min
Response: 118868
Conc: 1.53 ug/ml

Instrument :
FID_C
ClientSampleId :




#20 n-Hexatriacontane (C36)

R.T.: 20.126 min
Delta R.T.: -0.005 min
Response: 110660
Conc: 1.54 ug/ml



#21 n-Octatriacontane (C38)

R.T.: 20.868 min
Delta R.T.: -0.009 min
Response: 125560
Conc: 1.83 ug/ml



#22 n-Tetracontane (C40)

R.T.: 21.813 min
Delta R.T.: 0.011 min
Response: 490315
Conc: 6.82 ug/ml