

rters

Instrument :
FID_C
LabSampleId :
20 PPM ALIPHATIC HC STD
Area Percent Report

Manual IntegrationsAPPROVED

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC03102
Data File : FC058669.D
Signal(s) : FID1A.CH
Acq On : 10 Mar 2022 2:25 pm
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 2 Sample Multiplier: 1

Reviewed By :Yogesh Patel 03/11/2022
Supervised By :Ankita Jodhani 03/11/2022

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 022322.M
Title : GC Extractables

Signal : FID1A.CH

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.250	3.225	3.338	BB	283429	2392157	92.25%	4.846%
2	4.300	4.268	4.385	BB	264207	2380200	91.79%	4.822%
3	5.862	5.835	6.011	BB	237376	2526904	97.45%	5.119%
4	6.295	6.248	6.371	BB	245394	2454226	94.64%	4.972%
5	6.907	6.868	7.058	BB	234372	2529244	97.54%	5.124%
6	8.073	8.038	8.158	BB	232523	2468202	95.18%	5.000%
7	9.659	9.618	9.741	BB	224421	2459883	94.86%	4.983%
8	11.090	10.981	11.161	BB	211728	2434963	93.90%	4.933%
9	11.396	11.358	11.475	BB	244226	2593132	100.00%	5.253%
10	12.390	12.351	12.458	BB	194571	2352838	90.73%	4.766%
11	12.828	12.791	12.898	BB	163319	1970549	75.99%	3.992%
12	12.997	12.961	13.085	BB	187630	2340908	90.27%	4.742%
13	13.579	13.501	13.648	BB	183464	2290489	88.33%	4.640%
14	14.674	14.595	14.738	BB	181867	2207504	85.13%	4.472%
15	15.689	15.618	15.758	BB	168349	2136804	82.40%	4.329%
16	16.635	16.521	16.688	BV	158749	2094226	80.76%	4.243%
17	17.516	17.435	17.575	PB	156166	2087340	80.49%	4.229%
18	18.345	18.278	18.393	M	146523	2031429	78.34%	4.115%
19	19.125	19.081	19.170	BV	133702	1915453	73.87%	3.880%
20	19.861	19.811	19.894	BV	132282	2020526	77.92%	4.093%
21	20.565	20.511	20.675	BV	111841	1817827	70.10%	3.683%
22	21.391	21.328	21.460	BBA	85582	1857733	71.64%	3.763%
Sum of corrected areas:						49362536		

Aliphatic EPH 022322.M Fri Mar 11 09:30:33 2022