

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC031023AL\
 Data File : FC063147.D
 Signal(s) : FID1A.ch
 Acq On : 10 Mar 2023 12:22
 Operator : YP/AJ
 Sample : 20 PPM ALIPHATIC HC STD
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Integration File: autoint1.e
 Quant Time: Mar 11 03:17:53 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 022223.M
 Quant Title : GC Extractables
 QLast Update : Wed Feb 22 04:36:00 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.535	2560168	19.724 ug/ml
Spiked Amount 50.000		Recovery =	39.45%
12) S 1-chlorooctadecane (S...	12.972	1935096	19.837 ug/ml
Spiked Amount 50.000		Recovery =	39.67%
Target Compounds			
1) T n-Nonane (C9)	3.296	2346259	19.790 ug/ml
2) T n-Decane (C10)	4.364	2295220	19.904 ug/ml
3) T A~Naphthalene (C11.7)	5.948	2547660	19.921 ug/ml
4) T n-Dodecane (C12)	6.383	2297845	19.899 ug/ml
5) T A~2-methylnaphthalene...	7.005	2348745	19.955 ug/ml
6) T n-Tetradecane (C14)	8.178	2259795	19.995 ug/ml
7) T n-Hexadecane (C16)	9.779	2291335	19.918 ug/ml
8) T n-Octadecane (C18)	11.222	2328768	19.994 ug/ml
10) T n-Eicosane (C20)	12.530	2269180	19.915 ug/ml
11) T n-Heneicosane (C21)	13.141	2270834	19.926 ug/ml
13) T n-Docosane (C22)	13.726	2253363	19.980 ug/ml
14) T n-Tetracosane (C24)	14.827	2262246	20.187 ug/ml
15) T n-Hexacosane (C26)	15.848	2263554	20.468 ug/ml
16) T n-Octacosane (C28)	16.796	2279150	20.714 ug/ml
17) T n-Tricontane (C30)	17.681	2337241	21.224 ug/ml
18) T n-Dotriacontane (C32)	18.513	2387015	21.388 ug/ml
19) T n-Tetratriacontane (C34)	19.295	2347584	21.005 ug/ml
20) T n-Hexatriacontane (C36)	20.033	2255555	20.778 ug/ml
21) T n-Octatriacontane (C38)	20.762	2236392	20.580 ug/ml
22) T n-Tetracontane (C40)	21.660	2235382	20.547 ug/ml

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC031023AL\
Data File : FC063147.D
Signal(s) : FID1A.ch
Acq On : 10 Mar 2023 12:22
Operator : YP/AJ
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File: autoint1.e
Quant Time: Mar 11 03:17:53 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 022223.M
Quant Title : GC Extractables
QLast Update : Wed Feb 22 04:36:00 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um

