

rteres

Instrument :
FID_C
LabSampleId :
20 PPM ALIPHATIC HC STD
Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 03/14/2022
Supervised By :Ankita Jodhani 03/14/2022

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC03112
Data File : FC058705.D
Signal(s) : FID1A.CH
Acq On : 11 Mar 2022 22:09
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 2 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 022322.M
Title : GC Extractables

Signal : FID1A.CH

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.253	3.228	3.350	BBA	286618	2428373	89.84%	4.747%
2	4.302	4.275	4.391	BB	260835	2415734	89.38%	4.723%
3	5.864	5.818	6.021	BB	236401	2555094	94.53%	4.995%
4	6.295	6.265	6.388	BB	249553	2502972	92.60%	4.893%
5	6.908	6.795	7.075	BB	227185	2566545	94.96%	5.018%
6	8.073	8.035	8.165	BB	242760	2516972	93.12%	4.921%
7	9.661	9.555	9.745	BB	233622	2533440	93.73%	4.953%
8	11.091	11.045	11.165	BB	214573	2518206	93.17%	4.923%
9	11.398	11.361	11.478	BB	252842	2702891	100.00%	5.284%
10	12.391	12.355	12.468	BB	200419	2462322	91.10%	4.814%
11	12.831	12.795	12.908	BB	173446	2072953	76.69%	4.053%
12	13.000	12.961	13.068	BB	200035	2444164	90.43%	4.778%
13	13.582	13.515	13.655	BB	206026	2420466	89.55%	4.732%
14	14.678	14.581	14.738	BB	190015	2351630	87.00%	4.597%
15	15.693	15.485	15.761	BB	174970	2286194	84.58%	4.469%
16	16.638	16.528	16.705	BB	164750	2233711	82.64%	4.367%
17	17.520	17.448	17.585	BB	161269	2221478	82.19%	4.343%
18	18.349	18.028	18.421	PV	148681	2169515	80.27%	4.241%
19	19.129	19.081	19.191	BV	130099	2007869	74.29%	3.925%
20	19.865	19.811	19.910	BV	137437	2050343	75.86%	4.008%
21	20.572	20.511	20.658	BV	114415	1814699	67.14%	3.548%
22	21.398	21.328	21.460	BBA	84350	1875926	69.40%	3.667%
Sum of corrected areas:						51151497		

Aliphatic EPH 022322.M Mon Mar 14 08:52:53 2022