

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC031221AL\
 Data File : FC052389.D
 Signal(s) : FID1A.CH
 Acq On : 12 Mar 2021 10:16
 Operator : DD\AJ
 Sample : 20 PPM ALIPHATIC HC STD
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 FID_C
 LabSampleId :
 20 PPM ALIPHATIC HC STD

Integration File: autoint1.e
 Quant Time: Mar 13 02:16:51 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Alaphatic EPH 030321.M
 Quant Title : GC Extractables
 QLast Update : Thu Mar 04 03:56:17 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 T n-Nonane (C9)	117.875	117.857 E3	0.0	97	0.00
2 T n-Decane (C10)	118.765	119.836 E3	-0.9	98	0.00
3 T A~Naphthalene (C11.7)	132.777	133.705 E3	-0.7	98	0.00
4 T n-Dodecane (C12)	120.889	122.059 E3	-1.0	98	0.00
5 T A~2-methylnaphthalene (C12.	135.734	135.474 E3	0.2	97	0.00
6 T n-Tetradecane (C14)	121.243	121.721 E3	-0.4	98	0.00
7 T n-Hexadecane (C16)	119.650	118.264 E3	1.2	96	0.00
8 T n-Octadecane (C18)	123.787	119.659 E3	3.3	94	0.00
9 S ortho-Terphenyl (SURR)	131.876	121.858 E3	7.6	90	0.00
10 T n-Eicosane (C20)	123.218	115.403 E3	6.3	91	0.00
11 T n-Heneicosane (C21)	122.073	112.463 E3	7.9	90	0.00
12 S 1-chlorooctadecane (SURR)	104.000	95.035 E3	8.6	89	0.00
13 T n-Docosane (C22)	120.322	109.502 E3	9.0	89	0.00
14 T n-Tetracosane (C24)	116.584	104.039 E3	10.8	87	0.00
15 T n-Hexacosane (C26)	109.853	97.762 E3	11.0	87	0.00
16 T n-Octacosane (C28)	102.500	94.029 E3	8.3	90	0.00
17 T n-Tricontane (C30)	94.140	91.281 E3	3.0	96	0.00
18 T n-Dotriacontane (C32)	85.382	86.681 E3	-1.5	102	0.00
19 T n-Tetratriacontane (C34)	77.528	82.539 E3	-6.5	106	0.00
20 T n-Hexatriacontane (C36)	71.857	78.820 E3	-9.7	108	0.00
21 T n-Octatriacontane (C38)	68.501	74.964 E3	-9.4	108	-0.01
22 T n-Tetracontane (C40)	71.910	76.967 E3	-7.0	104	-0.01

Evaluate Continuing Calibration Report - Not Found

(#) = Out of Range

SPCC's out = 0 CCC's out = 0