

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC032124AL\
Data File : FC066091.D
Signal(s) : FID1A.ch
Acq On : 22 Mar 2024 03:04
Operator : YP/AJ
Sample : P1820-01
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :
WASTE-WATER

Integration File: autoint1.e
Quant Time: Mar 22 04:51:30 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 030424.M
Quant Title : GC Extractables
QLast Update : Mon Mar 04 13:09:51 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
12) S 1-chlorooctadecane (S...	12.898	3515398	24.113 ug/ml
Spiked Amount 50.000		Recovery =	48.23%

Target Compounds

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC032124AL\
Data File : FC066091.D
Signal(s) : FID1A.ch
Acq On : 22 Mar 2024 03:04
Operator : YP/AJ
Sample : P1820-01
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :
WASTE-WATER

Integration File: autoint1.e
Quant Time: Mar 22 04:51:30 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 030424.M
Quant Title : GC Extractables
QLast Update : Mon Mar 04 13:09:51 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 0.000 min
Exp R.T.: 11.493 min
Response: 0
Conc: N.D.

Instrument :
FID_C
ClientSampleId :
WASTE-WATER



#12 1-chlorooctadecane (SURR)

R.T.: 12.898 min
Delta R.T.: -0.026 min
Response: 3515398
Conc: 24.11 ug/ml