

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC032221AL\
 Data File : FC052616.D
 Signal(s) : FID1A.CH
 Acq On : 22 Mar 2021 18:39
 Operator : DD\AJ
 Sample : M1721-02
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :

Integration File: autoint1.e
 Quant Time: Mar 23 01:44:06 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Alaphatic EPH 031721.M
 Quant Title : GC Extractables
 QLast Update : Wed Mar 17 03:47:28 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.566	4897694	36.539 ug/ml
Spiked Amount 50.000		Recovery =	73.08%
12) S 1-chlorooctadecane (S...	13.004	3898306	36.143 ug/ml
Spiked Amount 50.000		Recovery =	72.29%
Target Compounds			
6) T n-Tetradecane (C14)	8.264f	61541	0.509 ug/ml
10) T n-Eicosane (C20)	12.611	280759	2.199 ug/ml
14) T n-Tetracosane (C24)	14.827	124919	1.000 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC032221AL\
Data File : FC052616.D
Signal(s) : FID1A.CH
Acq On : 22 Mar 2021 18:39
Operator : DD\AJ
Sample : M1721-02
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :

Integration File: autoint1.e
Quant Time: Mar 23 01:44:06 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Alaphatic EPH 031721.M
Quant Title : GC Extractables
QLast Update : Wed Mar 17 03:47:28 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#6 n-Tetradecane (C14)


R.T.: 8.264 min
Delta R.T.: 0.053 min
Response: 61541
Conc: 0.51 ug/ml

Instrument :
FID_C
ClientSampleId :




#9 ortho-Terphenyl (SURR)

R.T.: 11.566 min
Delta R.T.: -0.002 min
Response: 4897694
Conc: 36.54 ug/ml




#10 n-Eicosane (C20)

R.T.: 12.611 min
Delta R.T.: 0.050 min
Response: 280759
Conc: 2.20 ug/ml



#12 1-chlorooctadecane (SURR)

R.T.: 13.004 min
Delta R.T.: -0.002 min
Response: 3898306
Conc: 36.14 ug/ml



#14 n-Tetracosane (C24)

R.T.: 14.827 min
Delta R.T.: -0.032 min
Response: 124919
Conc: 1.00 ug/ml

Instrument :
FID_C
ClientSampleId :