

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC032221AL\
 Data File : FC052621.D
 Signal(s) : FID1A.CH
 Acq On : 22 Mar 2021 21:48
 Operator : DD\AJ
 Sample : M1723-04
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :

Integration File: autoint1.e
 Quant Time: Mar 23 01:45:23 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Alaphatic EPH 031721.M
 Quant Title : GC Extractables
 QLast Update : Wed Mar 17 03:47:28 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.566	5243121	39.116 ug/ml
Spiked Amount 50.000		Recovery =	78.23%
12) S 1-chlorooctadecane (S...	13.003	4142065	38.403 ug/ml
Spiked Amount 50.000		Recovery =	76.81%
Target Compounds			
1) T n-Nonane (C9)	3.325	5906247	50.851 ug/ml
2) T n-Decane (C10)	4.394	6483835	55.330 ug/ml
3) T A~Naphthalene (C11.7)	5.977	569907	4.353 ug/ml
4) T n-Dodecane (C12)	6.409	665817	5.545 ug/ml
5) T A~2-methylnaphthalene...	7.012	80902	0.604 ug/ml
6) T n-Tetradecane (C14)	8.205	765593	6.330 ug/ml
10) T n-Eicosane (C20)	12.611f	96136	0.753 ug/ml

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC032221AL\
Data File : FC052621.D
Signal(s) : FID1A.CH
Acq On : 22 Mar 2021 21:48
Operator : DD\AJ
Sample : M1723-04
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :

Integration File: autoint1.e
Quant Time: Mar 23 01:45:23 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Alaphatic EPH 031721.M
Quant Title : GC Extractables
QLast Update : Wed Mar 17 03:47:28 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 uL
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#1 n-Nonane (C9)


R.T.: 3.325 min
Delta R.T.: -0.005 min
Response: 5906247
Conc: 50.85 ug/ml

Instrument :
FID_C
ClientSampleId :




#2 n-Decane (C10)

R.T.: 4.394 min
Delta R.T.: -0.002 min
Response: 6483835
Conc: 55.33 ug/ml




#3 A~Naphthalene (C11.7)

R.T.: 5.977 min
Delta R.T.: -0.005 min
Response: 569907
Conc: 4.35 ug/ml



#4 n-Dodecane (C12)


R.T.: 6.409 min
Delta R.T.: -0.006 min
Response: 665817
Conc: 5.55 ug/ml



#5 A~2-methylnaphthalene (C12.89)


R.T.: 7.012 min
Delta R.T.: -0.027 min
Response: 80902
Conc: 0.60 ug/ml

Instrument :
FID_C
ClientSampleId :




#6 n-Tetradecane (C14)

R.T.: 8.205 min
Delta R.T.: -0.006 min
Response: 765593
Conc: 6.33 ug/ml




#9 ortho-Terphenyl (SURR)

R.T.: 11.566 min
Delta R.T.: -0.002 min
Response: 5243121
Conc: 39.12 ug/ml



#10 n-Eicosane (C20)

R.T.: 12.611 min
Delta R.T.: 0.050 min
Response: 96136
Conc: 0.75 ug/ml



#12 1-chlorooctadecane (SURR)

R.T.: 13.003 min
Delta R.T.: -0.003 min
Response: 4142065
Conc: 38.40 ug/ml

Instrument :
FID_C
ClientSampleId :