

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC032223AL\
 Data File : FC063207.D
 Signal(s) : FID1A.ch
 Acq On : 22 Mar 2023 10:00
 Operator : YP/AJ
 Sample : 02020-13
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :
 SP-2A

Integration File: autoint1.e
 Quant Time: Mar 23 00:19:09 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 031523.M
 Quant Title : GC Extractables
 QLast Update : Thu Mar 16 01:30:46 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.527	3055599	23.339 ug/ml
Spiked Amount 50.000		Recovery =	46.68%
12) S 1-chlorooctadecane (S...	12.967	2314166	23.984 ug/ml
Spiked Amount 50.000		Recovery =	47.97%

Target Compounds

(f)=RT Delta > 1/2 Window


(m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC032223AL\
Data File : FC063207.D
Signal(s) : FID1A.ch
Acq On : 22 Mar 2023 10:00
Operator : YP/AJ
Sample : 02020-13
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
FID_C
ClientSampleId :
SP-2A

Integration File: autoint1.e
Quant Time: Mar 23 00:19:09 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 031523.M
Quant Title : GC Extractables
QLast Update : Thu Mar 16 01:30:46 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.527 min
Delta R.T.: -0.005 min
Response: 3055599
Conc: 23.34 ug/ml

Instrument :
FID_C
ClientSampleId :
SP-2A



#12 1-chlorooctadecane (SURR)

R.T.: 12.967 min
Delta R.T.: -0.004 min
Response: 2314166
Conc: 23.98 ug/ml