

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC032223AL\
 Data File : FC063213.D
 Signal(s) : FID1A.ch
 Acq On : 22 Mar 2023 15:07
 Operator : YP/AJ
 Sample : 02020-09DL 5X
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :
 SP-1GDL

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 03/23/2023
 Supervised By :Ankita Jodhani 03/23/2023

Integration File: autoint1.e
 Quant Time: Mar 23 00:19:50 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 031523.M
 Quant Title : GC Extractables
 QLast Update : Thu Mar 16 01:30:46 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.530	1377646	10.523 ug/mlm
Spiked Amount	50.000	Recovery	= 21.05%
12) S 1-chlorooctadecane (S...	12.968	1076430	11.156 ug/ml
Spiked Amount	50.000	Recovery	= 22.31%

Target Compounds

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC032223AL\
Data File : FC063213.D
Signal(s) : FID1A.ch
Acq On : 22 Mar 2023 15:07
Operator : YP/AJ
Sample : 02020-09DL 5X
Misc :
ALS Vial : 12 Sample Multiplier: 1


Instrument :
FID_C
ClientSampleId :
SP-1GDL


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/23/2023
Supervised By :Ankita Jodhani 03/23/2023

Integration File: autoint1.e
Quant Time: Mar 23 00:19:50 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 031523.M
Quant Title : GC Extractables
QLast Update : Thu Mar 16 01:30:46 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.530 min
Delta R.T.: -0.002 min
Response: 1377646
Conc: 10.52 ug/ml

Instrument :
FID_C
ClientSampleId :
SP-1GDL

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/23/2023
Supervised By :Ankita Jodhani 03/23/2023



#12 1-chlorooctadecane (SURR)

R.T.: 12.968 min
Delta R.T.: -0.003 min
Response: 1076430
Conc: 11.16 ug/ml