

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC032419AL\
 Data File : FC040079.D
 Signal(s) : FID1A.CH
 Acq On : 24 Mar 2019 11:56
 Operator : AJ\SJ
 Sample : 20 PPM ALIPHATIC HC STD
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 FID_C
 LabSampleId :
 20 PPM ALIPHATIC HC STD

Integration File: autoint1.e
 Quant Time: Mar 25 01:28:16 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Alaphatic EPH 032319.M
 Quant Title : GC Extractables
 QLast Update : Sat Mar 23 13:38:33 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound		AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 T	n-Nonane (C9)	82.533	82.431 E3	0.1	100	0.00
2 T	n-Decane (C10)	84.635	85.123 E3	-0.6	101	0.00
3 T	A~Naphthalene (C11.7)	97.569	97.691 E3	-0.1	101	0.00
4 T	n-Dodecane (C12)	86.377	87.224 E3	-1.0	102	0.00
5 T	A~2-methylnaphthalene (C12.	99.363	99.804 E3	-0.4	101	0.00
6 T	n-Tetradecane (C14)	91.142	92.230 E3	-1.2	103	0.00
7 T	n-Hexadecane (C16)	96.115	97.300 E3	-1.2	103	0.00
8 T	n-Octadecane (C18)	100.312	102.027 E3	-1.7	102	0.00
9 S	ortho-Terphenyl (SURR)	128.777	130.408 E3	-1.3	102	0.00
10 T	n-Eicosane (C20)	101.770	104.060 E3	-2.3	103	0.00
11 T	n-Heneicosane (C21)	100.842	103.410 E3	-2.5	103	0.00
12 S	1-chlorooctadecane (SURR)	98.780	101.639 E3	-2.9	103	0.00
13 T	n-Docosane (C22)	100.519	103.179 E3	-2.6	103	0.00
14 T	n-Tetracosane (C24)	99.563	102.380 E3	-2.8	103	0.00
15 T	n-Hexacosane (C26)	97.904	101.358 E3	-3.5	104	0.00
16 T	n-Octacosane (C28)	98.098	103.473 E3	-5.5	107	0.00
17 T	n-Tricontane (C30)	97.850	106.580 E3	-8.9	111	0.00
18 T	n-Dotriacontane (C32)	96.779	105.503 E3	-9.0	111	0.00
19 T	n-Tetratriacontane (C34)	83.829	91.972 E3	-9.7	111	0.00
20 T	n-Hexatriacontane (C36)	73.628	79.303 E3	-7.7	110	-0.02
21 T	n-Octatriacontane (C38)	69.099	74.158 E3	-7.3	109	-0.02
22 T	n-Tetracontane (C40)	67.786	72.208 E3	-6.5	107	-0.07

Evaluate Continuing Calibration Report - Not Found

(#) = Out of Range

SPCC's out = 0 CCC's out = 0