

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC032521AL\
 Data File : FC052731.D
 Signal(s) : FID1A.CH
 Acq On : 26 Mar 2021 00:00
 Operator : DD\AJ
 Sample : 20 PPM ALIPHATIC HC STD
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 FID_C
 LabSampleId :
 20 PPM ALIPHATIC HC STD

Integration File: autoint1.e
 Quant Time: Mar 26 01:08:09 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Alaphatic EPH 031721.M
 Quant Title : GC Extractables
 QLast Update : Wed Mar 17 03:47:28 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound		AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 T	n-Nonane (C9)	116.149	123.709 E3	-6.5	107	0.00
2 T	n-Decane (C10)	117.185	125.181 E3	-6.8	107	0.00
3 T	A~Naphthalene (C11.7)	130.926	138.317 E3	-5.6	106	0.00
4 T	n-Dodecane (C12)	120.066	127.716 E3	-6.4	106	0.00
5 T	A~2-methylnaphthalene (C12.	133.859	141.013 E3	-5.3	105	0.00
6 T	n-Tetradecane (C14)	120.946	126.840 E3	-4.9	105	0.00
7 T	n-Hexadecane (C16)	120.270	122.814 E3	-2.1	102	0.00
8 T	n-Octadecane (C18)	126.170	123.987 E3	1.7	98	0.00
9 S	ortho-Terphenyl (SURR)	134.041	126.205 E3	5.8	94	0.00
10 T	n-Eicosane (C20)	127.694	119.395 E3	6.5	94	0.00
11 T	n-Heneicosane (C21)	127.107	116.355 E3	8.5	91	0.00
12 S	1-chlorooctadecane (SURR)	107.859	98.243 E3	8.9	91	0.00
13 T	n-Docosane (C22)	126.680	112.976 E3	10.8	89	0.00
14 T	n-Tetracosane (C24)	124.921	107.007 E3	14.3	85	0.00
15 T	n-Hexacosane (C26)	120.412	101.033 E3	16.1	83	0.00
16 T	n-Octacosane (C28)	115.294	97.569 E3	15.4	84	0.00
17 T	n-Tricontane (C30)	108.950	96.049 E3	11.8	88	0.00
18 T	n-Dotriacontane (C32)	100.339	94.872 E3	5.4	95	0.00
19 T	n-Tetratriacontane (C34)	91.919	94.462 E3	-2.8	102	0.00
20 T	n-Hexatriacontane (C36)	91.188	93.359 E3	-2.4	101	0.00
21 T	n-Octatriacontane (C38)	81.615	90.456 E3	-10.8	110	0.00
22 T	n-Tetracontane (C40)	86.140	94.861 E3	-10.1	110	0.00

Evaluate Continuing Calibration Report - Not Found

(#) = Out of Range

SPCC's out = 0 CCC's out = 0