

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC032621AL\
 Data File : FC052776.D
 Signal(s) : FID1A.CH
 Acq On : 27 Mar 2021 4:52
 Operator : DD\AJ
 Sample : 20 PPM ALIPHATIC HC STD
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 FID_C
 LabSampleId :
 20 PPM ALIPHATIC HC STD

Integration File: autoint1.e
 Quant Time: Mar 27 05:41:08 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Alaphatic EPH 031721.M
 Quant Title : GC Extractables
 QLast Update : Wed Mar 17 03:47:28 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 T n-Nonane (C9)	116.149	122.534 E3	-5.5	106	0.00
2 T n-Decane (C10)	117.185	124.927 E3	-6.6	106	0.00
3 T A~Naphthalene (C11.7)	130.926	137.870 E3	-5.3	105	0.00
4 T n-Dodecane (C12)	120.066	127.786 E3	-6.4	106	0.00
5 T A~2-methylnaphthalene (C12.	133.859	140.350 E3	-4.8	104	0.00
6 T n-Tetradecane (C14)	120.946	126.728 E3	-4.8	104	0.00
7 T n-Hexadecane (C16)	120.270	123.164 E3	-2.4	102	0.00
8 T n-Octadecane (C18)	126.170	124.452 E3	1.4	99	0.00
9 S ortho-Terphenyl (SURR)	134.041	126.367 E3	5.7	94	0.00
10 T n-Eicosane (C20)	127.694	119.936 E3	6.1	94	0.00
11 T n-Heneicosane (C21)	127.107	116.210 E3	8.6	91	0.00
12 S 1-chlorooctadecane (SURR)	107.859	98.378 E3	8.8	91	0.00
13 T n-Docosane (C22)	126.680	113.502 E3	10.4	89	0.00
14 T n-Tetracosane (C24)	124.921	107.753 E3	13.7	86	0.00
15 T n-Hexacosane (C26)	120.412	102.669 E3	14.7	85	0.00
16 T n-Octacosane (C28)	115.294	99.716 E3	13.5	86	0.00
17 T n-Tricontane (C30)	108.950	99.334 E3	8.8	91	0.00
18 T n-Dotriacontane (C32)	100.339	97.872 E3	2.5	98	0.00
19 T n-Tetratriacontane (C34)	91.919	97.265 E3	-5.8	105	0.00
20 T n-Hexatriacontane (C36)	91.188	95.545 E3	-4.8	103	0.00
21 T n-Octatriacontane (C38)	81.615	92.736 E3	-13.6	113	0.00
22 T n-Tetracontane (C40)	86.140	95.932 E3	-11.4	111	0.00

Evaluate Continuing Calibration Report - Not Found

(#) = Out of Range

SPCC's out = 0 CCC's out = 0