

rteres

Instrument :
FID_C
LabSampleId :
20 PPM ALIPHATIC HC STD
Area Percent Report

Manual IntegrationsAPPROVED

Reviewed By :Yogesh Patel 03/27/2025
Supervised By :mohammad ahmed 03/28/2025

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC032625
Data File : FC068415.D
Signal(s) : FID1A.ch
Acq On : 26 Mar 2025 20:24
Sample : 20 PPM ALIPHATIC HC STD
Misc :
ALS Vial : 16 Sample Multiplier: 1

Integration File: autoint1.e

Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 032625.M
Title : GC Extractables

Signal : FID1A.ch

peak #	R.T. min	Start min	End min	PK TY	peak height	peak area	peak % max.	% of total
1	3.198	3.171	3.261	BB	393543	3264585	79.87%	4.524%
2	4.254	4.221	4.314	BB	359749	3314400	81.09%	4.593%
3	5.819	5.781	5.921	BB	375946	3631109	88.84%	5.032%
4	6.261	6.221	6.318	BB	337043	3370500	82.46%	4.671%
5	6.870	6.831	6.978	BB	339086	3510124	85.88%	4.865%
6	8.048	8.011	8.134	BB	317918	3388430	82.90%	4.696%
7	9.643	9.608	9.731	BB	305079	3547172	86.78%	4.916%
8	11.081	11.051	11.154	BB	307155	3622672	88.63%	5.021%
9	11.385	11.341	11.458	BB	363451	4087411	100.00%	5.665%
10	12.386	12.348	12.448	BB	292715	3517926	86.07%	4.875%
11	12.826	12.768	12.908	BB	265575	3076360	75.26%	4.264%
12	12.996	12.954	13.071	BB	277261	3454961	84.53%	4.788%
13	13.581	13.511	13.654	BB	284664	3454716	84.52%	4.788%
14	14.681	14.601	14.754	BB	263559	3429307	83.90%	4.753%
15	15.699	15.634	15.774	BB	241540	3351136	81.99%	4.644%
16	16.647	16.601	16.699	BBA	241525	3320200	81.23%	4.601%
17	17.533	17.454	17.608	BB	248383	3491579	85.42%	4.839%
18	18.364	18.311	18.431	BB	237078	3409748	83.42%	4.726%
19	19.146	19.101	19.221	BB	206460	3033993	74.23%	4.205%
20	19.885	19.838	19.922	BV	165866	2539370	62.13%	3.519%
21	20.596	20.551	20.658	BB	136441	2241356	54.84%	3.106%
22	21.442	21.378	21.551	BB	90884	2098305	51.34%	2.908%
Sum of corrected areas:						7215359		

Aliphatic EPH 032625.M Thu Mar 27 02:07:35 2025