

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC032922AL\
 Data File : FC058970.D
 Signal(s) : FID1A.CH
 Acq On : 30 Mar 2022 8:50
 Operator : UA/NP
 Sample : N2149-04D
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :
 PILE-LINDEN-PROPANE-EPH-2

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 03/30/2022
 Supervised By :mohammad ahmed 04/01/2022

Integration File: sample.E
 Quant Time: Mar 30 09:18:43 2022
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 032522.M
 Quant Title : GC Extractables
 QLast Update : Fri Mar 25 23:18:46 2022
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.095	5851748	40.614 ug/ml
Spiked Amount	50.000	Recovery	= 81.23%
12) S 1-chlorooctadecane (S...	12.543	4281537	38.196 ug/mlm
Spiked Amount	50.000	Recovery	= 76.39%

Target Compounds

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC032922AL\
Data File : FC058970.D
Signal(s) : FID1A.CH
Acq On : 30 Mar 2022 8:50
Operator : UA/NP
Sample : N2149-04D
Misc :
ALS Vial : 26 Sample Multiplier: 1


Instrument :
FID_C
ClientSampleId :
PILE-LINDEN-PROPANE-EPH-2


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/30/2022
Supervised By :mohammad ahmed 04/01/2022

Integration File: sample.E
Quant Time: Mar 30 09:18:43 2022
Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 032522.M
Quant Title : GC Extractables
QLast Update : Fri Mar 25 23:18:46 2022
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
Signal Phase : Rxi-1ms
Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.095 min
Delta R.T.: -0.004 min
Response: 5851748
Conc: 40.61 ug/ml

Instrument :
FID_C
ClientSampleId :
PILE-LINDEN-PROPANE-EPH-2

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 03/30/2022
Supervised By :mohammad ahmed 04/01/2022



#12 1-chlorooctadecane (SURR)

R.T.: 12.543 min
Delta R.T.: -0.005 min
Response: 4281537
Conc: 38.20 ug/ml m