

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC033019AL\
 Data File : FC040230.D
 Signal(s) : FID1A.CH
 Acq On : 30 Mar 2019 14:11
 Operator : AJ\SJ
 Sample : 20 PPM ALIPHATIC HC STD
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 FID_C
 LabSampleId :
 20 PPM ALIPHATIC HC STD

Integration File: autoint1.e
 Quant Time: Apr 01 01:23:52 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Alaphatic EPH 032319.M
 Quant Title : GC Extractables
 QLast Update : Sat Mar 23 13:38:33 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Large solvent peaks clipped

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 T n-Nonane (C9)	82.533	83.462 E3	-1.1	101	0.00
2 T n-Decane (C10)	84.635	87.513 E3	-3.4	104	0.00
3 T A~Naphthalene (C11.7)	97.569	105.928 E3	-8.6	109	0.00
4 T n-Dodecane (C12)	86.377	92.273 E3	-6.8	108	0.00
5 T A~2-methylnaphthalene (C12.	99.363	108.928 E3	-9.6	111	0.00
6 T n-Tetradecane (C14)	91.142	97.818 E3	-7.3	109	0.00
7 T n-Hexadecane (C16)	96.115	100.374 E3	-4.4	106	0.00
8 T n-Octadecane (C18)	100.312	100.844 E3	-0.5	101	0.00
9 S ortho-Terphenyl (SURR)	128.777	130.586 E3	-1.4	102	0.01
10 T n-Eicosane (C20)	101.770	101.138 E3	0.6	100	0.00
11 T n-Heneicosane (C21)	100.842	100.461 E3	0.4	100	0.00
12 S 1-chlorooctadecane (SURR)	98.780	100.031 E3	-1.3	101	0.00
13 T n-Docosane (C22)	100.519	100.433 E3	0.1	100	0.00
14 T n-Tetracosane (C24)	99.563	100.561 E3	-1.0	101	0.01
15 T n-Hexacosane (C26)	97.904	100.995 E3	-3.2	104	0.01
16 T n-Octacosane (C28)	98.098	103.279 E3	-5.3	107	0.01
17 T n-Tricontane (C30)	97.850	105.020 E3	-7.3	110	0.01
18 T n-Dotriacontane (C32)	96.779	102.305 E3	-5.7	108	0.02
19 T n-Tetratriacontane (C34)	83.829	89.665 E3	-7.0	109	0.02
20 T n-Hexatriacontane (C36)	73.628	78.302 E3	-6.3	108	0.02
21 T n-Octatriacontane (C38)	69.099	72.320 E3	-4.7	106	0.03
22 T n-Tetracontane (C40)	67.786	69.354 E3	-2.3	103	0.01

Evaluate Continuing Calibration Report - Not Found

(#) = Out of Range

SPCC's out = 0 CCC's out = 0