

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC040723AL\
 Data File : FC063339.D
 Signal(s) : FID1A.ch
 Acq On : 07 Apr 2023 19:42
 Operator : YP/AJ
 Sample : 02230-01DL 5X
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 FID_C
 ClientSampleId :
 OK-01-040623DL

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/10/2023
 Supervised By :Ankita Jodhani 04/10/2023

Integration File: autoint1.e
 Quant Time: Apr 10 05:50:37 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 040523.M
 Quant Title : GC Extractables
 QLast Update : Wed Apr 05 07:22:46 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 ul
 Signal Phase : Rxi-1ms
 Signal Info : 20M x 0.18mm x 0.18um

Compound	R.T.	Response	Conc Units

System Monitoring Compounds			
9) S ortho-Terphenyl (SURR)	11.532	1132483	7.443 ug/ml
Spiked Amount 50.000		Recovery =	14.89%
12) S 1-chlorooctadecane (S...	12.970	962836	8.702 ug/mlm
Spiked Amount 50.000		Recovery =	17.40%

Target Compounds

(f)=RT Delta > 1/2 Window

(m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\FID_C\Data\FC040723AL\
Data File : FC063339.D
Signal(s) : FID1A.ch
Acq On : 07 Apr 2023 19:42
Operator : YP/AJ
Sample : 02230-01DL 5X
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :

FID_C

Client Sample Id :

OK-01-040623DL

Manual Integrations

APPROVED

Reviewed By : Yogesh Patel 04/10/2023

Supervised By : Ankita Jodhani 04/10/2023

Integration File: autoint1.e

Quant Time: Apr 10 05:50:37 2023

Quant Method : Z:\pestpcbsrv\HPCHEM1\FID_C\Method\Aliphatic EPH 040523.M

Quant Title : GC Extractables

QLast Update : Wed Apr 05 07:22:46 2023


Response via : Initial Calibration


Integrator: ChemStation

Volume Inj. : 1 ul

Signal Phase : Rxi-1ms

Signal Info : 20M x 0.18mm x 0.18um






#9 ortho-Terphenyl (SURR)

R.T.: 11.532 min
Delta R.T.: 0.005 min
Response: 1132483
Conc: 7.44 ug/ml

Instrument :
FID_C
ClientSampleId :
OK-01-040623DL

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/10/2023
Supervised By :Ankita Jodhani 04/10/2023



#12 1-chlorooctadecane (SURR)

R.T.: 12.970 min
Delta R.T.: 0.003 min
Response: 962836
Conc: 8.70 ug/ml m